

Measuring Module Digital



6.02100.010

Product manual

8.0108.8013EN / 2021-07-23



Metrohm AG
Ionenstrasse
CH-9100 Herisau
Switzerland
+41 71 353 85 85
info@metrohm.com
www.metrohm.com

Measuring Module Digital

6.02100.010

Product manual

8.0108.8013EN /
2021-07-23

Technical Communication
Metrohm AG
CH-9100 Herisau

This documentation is protected by copyright. All rights reserved.

This documentation has been prepared with great care. However, errors can never be entirely ruled out. Please send comments regarding possible errors to the address above.

Disclaimer

Deficiencies arising from circumstances that are not the responsibility of Metrohm, such as improper storage or improper use, etc., are expressly excluded from the warranty. Unauthorized modifications to the product (e.g. conversions or attachments) exclude any liability on the part of the manufacturer for resulting damage and its consequences. Instructions and notes in the Metrohm product documentation must be strictly followed. Otherwise, Metrohm's liability is excluded.

Table of contents

1	Overview	1
1.1	Measuring Module Digital – Product description	1
1.2	Measuring Module Digital – Product versions	1
1.3	Symbols and conventions	2
1.4	Further information	2
1.5	Accessories	2
2	Safety	4
2.1	Intended use	4
2.2	Responsibility of the operator	4
2.3	Requirements for operating personnel	5
2.4	Safety instructions	5
2.4.1	Danger from electrical potential	5
2.4.2	Danger from biological and chemical hazardous substances	5
2.4.3	Danger from highly flammable substances	6
2.5	Design of warning messages	6
2.6	Meaning of warning signs	7
3	Functional description	9
3.1	Measuring Module Digital – Overview	9
3.2	Measuring Module Digital – Functional description	9
3.3	System – Signals	10
3.4	Measuring Module Digital – Interfaces	10
4	Delivery and packaging	12
4.1	Delivery	12
4.2	Packaging	12
5	Installation	13
5.1	Mounting the measuring module	13
5.2	Plugging in the sensor	16
6	Start-up	18
6.1	Initial start-up by Metrohm	18



- 7 Operation and control 19**
 - 7.1 Operation 19
- 8 Maintenance 20**
 - 8.1 Maintenance 20
 - 8.2 Cleaning the product surface 20
- 9 Troubleshooting 22**
- 10 Disposal 23**
- 11 Technical specifications 24**
 - 11.1 Ambient conditions 24
 - 11.2 Measuring module – Energy supply 24
 - 11.3 Measuring module – Dimensions 24
 - 11.4 Measuring module – Housing 25
 - 11.5 Measuring Module Digital – Connector specifications 25
 - 11.6 Measuring Module Digital – Display specifications 25

1 Overview

1.1 Measuring Module Digital – Product description

The Measuring Module Digital is used as an interface for digital electrodes on an OMNIS Titrator or an OMNIS Titration Module.


1.2 Measuring Module Digital – Product versions

The product is available in the following versions:

Table 1 Product versions

Article number	Designation
6.02100.010	Measuring Module Digital

The article number and serial number for identifying the product can be found on the type plate:



1	(01) = Article number in accordance with GS1 standard	2	(21) = Serial number
3	(240) = Metrohm article number	4	Certification
5	Certification	6	Technical specifications



1.3 Symbols and conventions

The following formatting may appear in the documentation:

(5- 12)	Cross-reference to figure legend The first number refers to the figure number. The second number refers to the product part in the figure.
1	Instruction step Numbers indicate the order of the instructions steps.
Method	Names of parameters, menu items, tabs and dialog windows
File ► New	Menu path
[Continue]	Button or key

1.4 Further information

The Metrohm Knowledge Base <https://guide.metrohm.com> always provides the current version of this document. Further instructions, leaflets, release notes etc. may be available, depending on the product. You can directly access the required information or the associated PDF document using the full-text search function and filters.

1.5 Accessories

Up-to-date information on the scope of delivery and on optional accessories can be found on the Metrohm website. Download this information as follows:

Downloading the accessories list

- 1** Go to <https://www.metrohm.com>.
- 2** Enter the article number of the product (e.g. **2.1001.0010**) into the search field.

The search result is displayed.
- 3** Click on the product.

Detailed information regarding the product is shown on various tabs.

- 4 On the **Included parts** tab, click the link to download the PDF.
The PDF file with the accessories data is loaded.



NOTICE

Metrohm recommends downloading the accessories list from the Internet and keeping it for reference purposes.

2.3 Requirements for operating personnel

Only qualified personnel may operate the product. Qualified personnel are persons who meet the following requirements:

- Basic regulations on occupational safety and accident prevention for chemical laboratories are known and complied with.
- Knowledge of handling hazardous chemicals is present. Personnel have the ability to recognize and avoid potential dangers.
- Knowledge regarding the application of fire prevention measures for laboratories is available.
- Safety-relevant information is communicated and understood. The personnel can operate the product safely.
- The user documentation has been read and understood. The personnel operate the product according to the instructions in the user documentation.

2.4 Safety instructions

2.4.1 Danger from electrical potential

Contact with electrical potential can cause serious injuries or death. To avoid danger from electrical potential, observe the following:

- Operate the product only if it is in perfect condition. The housing must also be intact.
- Only use the product with the covers fitted. If covers are damaged or missing, disconnect the product from the energy supply and contact the regional Metrohm service representative.
- Protect live components (e.g. power supply unit, power cord, connection sockets) against moisture.
- Always have maintenance work and repairs on electrical components carried out by a regional Metrohm service representative.
- Disconnect the product from the energy supply immediately if at least one of the following cases occurs:
 - The housing is damaged or open.
 - Live parts are damaged.
 - Moisture penetrates.

2.4.2 Danger from biological and chemical hazardous substances

Contact with biological hazardous substances may cause poisoning from toxins or infections from microorganisms. Contact with aggressive chemical substances may cause poisoning or chemical burns. To avoid danger from biological or chemical hazardous substances, observe the following:



DANGER

Type and source of danger

Consequences when not observing the notice: An irreversible injury that may result in death is very probable.

- Measures to avoid the danger



WARNING

Type or source of danger

Consequences when not observing the notice: A serious injury that may result in death is probable.

- Measures to avoid the danger



CAUTION

Type or source of danger

Consequences when not observing the notice: A minor to moderate injury is probable.

- Measures to avoid the danger

2.6 Meaning of warning signs

This documentation uses the following warning signs:

Table 2 Warning sign according to ISO 7010

Warning sign	Meaning
	General warning sign
	Warning of electrical voltage
	Warning of hand injuries
	Warning of sharp object



Warning sign	Meaning
	Warning of hot surface
	Warning of biological hazard
	Warning of toxic materials
	Warning of flammable materials
	Warning of corrosive substances
	Warning of optical radiation
	Warning of laser beams

Depending on the intended use of the product, the corresponding warning sign stickers must be placed on the product.

3 Functional description

3.1 Measuring Module Digital – Overview



Figure 1 Measuring Module Digital – Overview

1 Fastening screws

Fastening screws, left and right. These fasten the measuring module in the housing and ground the electronics.

2 Connection socket

For dTodes

3.2 Measuring Module Digital – Functional description

The Measuring Module Digital is used as an interface for digital electrodes on an OMNIS Titrator or an OMNIS Titration Module.

The digital electrodes, the dTodes, can be plugged into its connection socket.

3.3 System – Signals

System components with status indicators show their operating status with colors and/or flashing patterns. The meaning of the colors and flashing patterns is explained in the following table.

Visual signal		Meaning
	LED lights up yellow.	System start or initialization
	LED flashes yellow (slowly).	Ready for connection setup or locking
	LED flashes yellow (fast).	Connection setup started or locking underway
	LED lights up green.	Ready for operation
	LED flashes green (slowly).	In operation
	LED flashes red (fast).	Malfunction or error

Some system components only use part of the explained flashing patterns.

3.4 Measuring Module Digital – Interfaces

The Measuring Module Digital has 1 connection socket for a digital electrode.



Figure 2 Measuring Module Digital – Interfaces and connectors

1 Connection socket for digital electrodes

Connection socket

The black marking around the connection socket indicates that only a cable of a digital electrode may be plugged in here.



4 Delivery and packaging

4.1 Delivery

Inspect the delivery immediately upon receipt:

- Check the delivery against the delivery note to ensure completeness.
- Check the product for damage.
- If the delivery is incomplete or damaged, contact your regional Metrohm representative.

4.2 Packaging


The product and accessories are supplied in protective special packaging. Keep this packaging to ensure safe transportation of the product. If a transport locking device is present, keep this as well for future reuse.

5 Installation

5.1 Mounting the measuring module


The measuring module is supplied with mounted fastening screws. These fastening screws are used to secure the measuring module in the instrument to ensure trouble-free operation.

1 Opening the lid



- Open the lid.

2 Removing the side parts




Carry out the following steps on **both sides** of the instrument.

- Remove the 2 screws from above using the hex key.




- Push the side covering upwards until it can be removed from the side.
- Remove the side covering to the side.

3 Removing the fastening screws




- Remove the 2 fastening screws from the measuring module using the hex key.

4 Inserting the measuring module




- Insert the measuring module into an empty slot. The slots are designated with 1 (rear) and 2 (front).

5 Attaching the measuring module



- Insert the fastening screws. Tighten the measuring module to the housing from **both sides** using the hex key.

6 Mounting the side parts




Carry out the following steps on **both sides** of the instrument.

- Position the side covering from the side in an elevated position.
- Insert the side covering into the guide rail and push it downwards.
- Insert the 2 screws and tighten them from above using the hex key.



5.2 Plugging in the sensor

1 Plugging in the electrode cable




NOTICE

It should be easy to plug in the plug.

- Do not apply force if the plug cannot be inserted easily.
- Rotate the plug to the right or left using light pressure until it latches in the socket.
- Align the red dot on the plug with the groove on the connection socket.
- Plug in the plug until you can feel it snap in.

2 Guiding out the cable



- Guide the cable out going under the bar.
- Close the lid.

See also

Measuring Module Digital – Interfaces (chapter 3.4, page 10)



6 Start-up

6.1 Initial start-up by Metrohm

As a basic rule, the initial start-up of the system is carried out by the regional Metrohm service representative.

7 Operation and control

7.1 Operation

The product can be operated via the OMNIS Software. Further information on the OMNIS Software under [OMNIS Help](#).



WARNING

Electrical potential

Contact with electrical potential can cause serious injuries or death.

- Operate the product only if it is in perfect condition. The housing must also be intact.
- Only use the product with the covers fitted.
- Protect live components (e.g. power supply unit, power cord, connection sockets) against moisture.
- Always have maintenance work and repairs on electrical components carried out by a regional Metrohm service representative.

Prerequisite:

- The product is switched off and disconnected from the energy supply.

Required accessories:

- Cleaning cloth (soft, lint-free)
- Water or ethanol

1 Clean the surface with a damp cloth. Remove persistent contamination with ethanol.

2 Wipe the surface with a dry cloth.

3 Clean the connectors with a dry cloth.

9 Troubleshooting

Messages on malfunctions and errors are displayed in the control software or in the embedded software (e.g. on the display of an instrument) and contain the following information:

- Descriptions of causes of malfunctions (e.g. jammed drive)
- Descriptions of problems with the control (e.g. missing or invalid parameter)
- Information on how to solve the problem


System components with status display elements also indicate malfunctions and errors with a red flashing LED.

Troubleshooting on the product is often only possible with the control software or the embedded software (e.g. initializing, moving to a defined position).

See also

System – Signals (chapter 3.3, page 10)

10 Disposal



Properly dispose of chemicals and of the product to reduce negative effects on the environment and public health. Local authorities, waste disposal companies or dealers provide more detailed information on disposal. Observe the WEEE EU directive (WEEE = Waste Electrical and Electronic Equipment) for the proper disposal of waste electronic equipment within the European Union.



11 Technical specifications

11.1 Ambient conditions

Nominal function range	+5 to +45 °C	at max. 80% relative humidity, non- condensing
Storage	+5 to +45 °C	

11.2 Measuring module – Energy supply

Power consumption	max. 0.6 W	-
Energy transmission	-	inductive coupling

11.3 Measuring module – Dimensions

Measurements

<i>Width</i>	105 mm
<i>Height</i>	31 mm
<i>Depth</i>	72 mm
Weight	approx. 420 g

11.4 Measuring module – Housing

Materials

<i>Lid</i>	AW-5754 H12 / H22	aluminum, coated
<i>Back panel</i>	PBT	poly(butylene terephthalate)
<i>Enclosure</i>	GD-ZnAl4Cu1	zinc die cast, nickel-plated

IP degree of protection IP 40

11.5 Measuring Module Digital – Connector specifications

Connection socket for a digital electrode

<i>Type</i>	round plug 6-pin, size 0, 60°
-------------	-------------------------------

11.6 Measuring Module Digital – Display specifications

Status display LED green