dAquatrode Plus

6.00202.300

Sensor leaflet

8.0109.8003EN / 2021-09-23

Metrohm

Metrohm AG
Ionenstrasse
CH-9100 Herisau
Switzerland
+41 71 353 85 85
info@metrohm.com
www.metrohm.com

dAquatrode Plus

6.00202.300

Sensor leaflet

8.0109.8003EN /
2021-09-23

Technical Communication
Metrohm AG
CH-9100 Herisau

This documentation is protected by copyright. All rights reserved.

This documentation has been prepared with great care. However, errors can never be entirely ruled out. Please send comments regarding possible errors to the address above.

Disclaimer

Deficiencies arising from circumstances that are not the responsibility of Metrohm, such as improper storage or improper use, etc., are expressly excluded from the warranty. Unauthorized modifications to the product (e.g. conversions or attachments) exclude any liability on the part of the manufacturer for resulting damage and its consequences. Instructions and notes in the Metrohm product documentation must be strictly followed. Otherwise, Metrohm's liability is excluded.

Table of contents

1 Overview 1

1.1 dAquatrode Plus – Product description ... 1

1.2 dAquatrode Plus – Overview ... 1

2 Functional description 2

2.1 pH electrode – Functional description ... 2

3 Delivery and packaging 3

3.1 Delivery ... 3

3.2 Packaging ... 3

3.3 Unpacking and assessing the electrode ... 3

3.4 Storing the dAquatrode Plus ... 4

4 Installation 5

4.1 Preparing the dAquatrode Plus ... 5

4.2 Mounting the electrode ... 6

5 Operation and control 8

5.1 Calibrating the pH electrode ... 8

6 Maintenance 9

6.1 dAquatrode Plus – Replacing the electrolyte ... 9

6.2 Cleaning the pH electrode ... 9

7 Troubleshooting 10

8 Electrode – Disposal 11

9 Technical specifications 12

9.1 Ambient conditions ... 12

9.2 pH electrode – Dimensions ... 12

9.3 pH electrode – Housing ... 12

9.4 pH electrode – Connectors specifications ... 12

9.5 dTrodes – Display specifications ... 13

9.6 dAquatrode Plus – Measurement specifications ... 13

1 Overview

1.1 dAquatrode Plus – Product description

The dAquatrode Plus is a combined pH glass electrode with Pt1000 temperature sensor for measurements in ion-deficient, aqueous solutions. The dAquatrode Plus is a dTrode (digital electrode) for OMNIS.

1.2 dAquatrode Plus – Overview



Figure 1 dAquatrode Plus

1	**Protective cap**	**2**	**Electrode head**
3	**Filler opening** For bridge electrolyte	**4**	**Fixed ground-joint diaphragm**

2 Functional description

2.1 pH electrode – Functional description

The glass membrane of the pH electrode consists of a silicate framework containing lithium ions. When the glass surface is immersed in an aqueous solution, then a thin hydrated layer (gel layer) forms on the outside and inside of the glass surface.

Because the proton concentration in the internal buffer of the pH electrode is constant (pH 7), a stationary condition arises on the internal side of the glass membrane. If the proton concentration in the measuring solution changes, an ion exchange takes place in the outer hydrated layer, thus leading to a change of potential on the glass membrane. Only when this ion exchange has reached a stationary condition, the potential of the pH electrode is also constant.

3 Delivery and packaging

3.1 Delivery

Inspect the delivery immediately upon receipt:

- Check the delivery against the delivery note to ensure completeness.
- Check the product for damage.
- If the delivery is incomplete or damaged, contact your regional Metrohm representative.

3.2 Packaging

The product and accessories are supplied in protective special packaging. Keep this packaging to ensure safe transportation of the product. If a transport locking device is present, keep this as well for future reuse.

3.3 Unpacking and assessing the electrode

1 Unpacking the electrode

Remove the electrode with storage vessel from the packaging.

2 Removing the storage vessel



Figure 2 Releasing the electrode from the storage vessel

- Hold the electrode and storage vessel firmly in your hand so that the electrode cannot slip away.
- Position the tool between the storage vessel and SGJ sleeve.

- **Carefully** push the tool to the side to release the electrode.
  Do not tip the tool forwards!

Avoid applying excess pressure to the tool. Otherwise, the electrode could be released too abruptly.

3 **Checking the electrode for proper function**

- **Preparing the electrode:**
  (see "Preparing the dAquatrode Plus", chapter 4.1, page 5)
- **Calibrating the electrode:**
  (see "Calibrating the pH electrode", chapter 5.1, page 8)

Defective electrodes must be sent back for warranty processing within 2 months (starting from the day of delivery).

3.4 Storing the dAquatrode Plus

⚠ CAUTION

Property damage caused by dried out sensor

Destruction of the sensor due to drying out.

- Do not let the sensor dry out.
- Follow the storage instructions.

The electrode head must be stored as follows to protect it from water, solvents, dust and mechanical influences:

1 Screw the protective cap *(1-**1**)* onto the electrode head *(1-**2**)*.

2 Store the electrode in the storage vessel. When doing so, ensure that the electrode is immersed in the storage solution via the fixed ground-joint diaphragm *(1-**4**)*.

We recommend storing electrodes with
3 mol/L of potassium chloride as a bridge electrolyte in the storage solution (6.2323.000). This prevents the glass membrane from aging and the electrode can be used without prior conditioning.

The storage solution may only be used for this electrolyte; we recommend to store all other electrolytes in the bridge electrolyte.

3 Close the filler opening *(1-**3**)*.

4 Installation

4.1 Preparing the dAquatrode Plus

1 Filling with bridge electrolyte

Open the closure of the filler opening *(1-**3**)* and, if necessary, fill the bridge electrolyte up to the filler opening.

2 Rinsing the electrode

⚠ CAUTION

Property damage caused by electrostatic charge

Useless measurement results due to electrostatically charged electrode and damage through mechanical treatment.

- Never dab the electrode membrane dry.

Rinse the electrode with distilled water.

3 Connecting the electrode

- Unscrew the protective cap *(1-**1**)*.
- Position the cable connection on the electrode head such that the slot in the cable connection is on the guide lug of the electrode head.
- Push the socket in the cable connection into the plug inside the electrode head.
- Push the outer ring of the cable connection over the electrode head.
  Ensure that the guide lugs in the electrode head are in the grooves of the cable connection.
- Push the cable connection onto the electrode head until it snaps in place.

ℹ To remove the cable, first release the outer ring and then carefully pull the cable connection from the electrode head.
When doing so, be sure not to pull on the cable itself but the cable connector instead.

4.2 Mounting the electrode

The electrode must sit securely in the titration head.

For automatic procedures, ensure that the cables have enough room to move.

During the titration, it is important that the solution is mixed well. The stirring rate should be high enough to form a small vortex. If the stirring rate is too high, then air bubbles will be aspirated. These may result in incorrect measured values. If the stirring rate is too low, then the solution is only mixed slowly and the reaction time or titration time increases accordingly.

In order for the measurement to be taken in a well-mixed solution after the addition of the titrant, the titration tip should be positioned where turbulence is high. Furthermore, the distance between the addition of the titrant and the electrode should be as large as possible. Therefore, take into account the stirring direction (counterclockwise or clockwise) when positioning the electrode and titration tip.



Figure 3 Diagrams showing rod stirrer, electrode and titration tip during a titration. a) clockwise stirring direction, b) counterclockwise stirring direction.

1	**Rod stirrer**	**2**	**Electrode**
3	**Titration tip**		

5 Operation and control

5.1 Calibrating the pH electrode

1 Rinse the electrode with distilled water.

2 **Calibrating the electrode with the first buffer**
Immerse the electrode in buffer solution (pH 7) and start the calibration.

3 After a successful measurement, remove the electrode from the buffer and rinse it with distilled water.

4 **Calibrating the electrode with the second buffer**
Repeat steps 2 and 3 with the second buffer.

5 **Calibrating the electrode with the third buffer, if necessary**
Repeat steps 2 and 3 with the third buffer.

6 Use the following information to determine if the electrode meets the requirements:

- **Slope:**
  95–103%
  pH:
  6.5–7.5
- **Offset of potential:**
  –30 to 30 mV

6 Maintenance

6.1 dAquatrode Plus – Replacing the electrolyte

1. Open the filler opening *(1-**3**)*.
2. Use a plastic pipette to drain the bridge electrolyte from the electrode.
3. Rinse the inside of the electrode with the new electrolyte and empty it again.
4. Fill the electrode with electrolyte up to the filler opening.
5. Close the filler opening *(1-**3**)* if the electrode is not used immediately.
6. Immerse the electrode in a storage solution overnight.

   Then, the electrode is ready for use again.

6.2 Cleaning the pH electrode

1. Never treat the electrode in an ultrasonic bath. The electrode could become damaged.

   Rinse the electrode with distilled water.

7 Troubleshooting

Problem	Cause	Remedy
Slow response	Electrostatic charge	Do not dab the glass membrane with a towel.
	Solid deposits on membrane surface	Clean the electrode with a solvent/strong acid.
	Deposition of proteins on membrane surface	Clean the electrode with 5% pepsin in 0.1 mol/L of HCl.
Zero point shift	The electrode was in dry storage	Store the electrode in a storage solution overnight.
	Solid deposits on membrane surface	Clean the electrode with a solvent/strong acid.
	Deposition of proteins on membrane surface	Clean the electrode with 5% pepsin in 0.1 mol/L of HCl.
	Reference system contaminated or dried out	Clean the reference system with a reference electrolyte and refill it. Then, store the electrode in storage solution.
	Ground-joint diaphragm contaminated	Clean the electrode with pHit kit (6.2325.000).
Slope too low	Solid deposits on membrane surface	Clean the electrode with a solvent/strong acid.
	Deposition of proteins on membrane surface	Clean the electrode with 5% pepsin in 0.1 mol/L of HCl.
	Reference system contaminated or dried out	Clean the reference system with a reference electrolyte and refill it. Then, store the electrode in storage solution.
	Ground-joint diaphragm contaminated	Clean the electrode with pHit kit (6.2325.000).

8 Electrode – Disposal



This product is covered by European Directive, WEEE – Waste Electrical and Electronic Equipment.

The correct disposal of your old instrument will help to prevent negative effects on the environment and public health.

Proceed as follows to dispose of the electrode:

1 Draining the electrolyte

Use a plastic pipette to remove the electrolyte from the electrode.

2 Disposing of the electrolyte

Dispose of the electrolyte in accordance with the legal provisions.

3 Disposing of the electrode

Put the electrode in electronic waste recycling.

More details about the disposal of your old product can be obtained from your local authorities, from waste disposal companies or from your local dealer.

9 Technical specifications

9.1 Ambient conditions

Nominal function range	+5 to +45 °C	at max. 80% relative humidity, non-condensing
Storage	+5 to +45 °C	

9.2 pH electrode – Dimensions

Measurements

Shaft diameter	12 mm
Maximum installation length	125 mm

9.3 pH electrode – Housing

Materials

Shaft material	Glass

9.4 pH electrode – Connectors specifications

Connector	Metrohm plug-in head Q

9.5 dTrodes – Display specifications

Status display	LED	green-red

9.6 dAquatrode Plus – Measurement specifications

pH range	0–13
Temperature range	0–60 °C
Minimum immersion depth	20 mm