



Sampling Solutions for NIRS Process Analyzers

Find the best solution for your sample

**PUSHING
THE
LIMITS
TOGETHER**

 **Metrohm**
Process Analytics

Monitoring Your Process 24/7

The right choice for success

Near-infrared (NIR) spectroscopic techniques are characterized by their ability to gain rapid, accurate information from the high-resolution spectra of solid, gas, and liquid samples without prior sample preparation or chemical reagents. They are economical and facilitate qualitative and quantitative as well as noninvasive and nondestructive analysis. For all these reasons, spectroscopic techniques are ideally suited for industrial quality control and process monitoring.

The determination of multiple parameters in a single measurement, and the possibility of fiber optics implementation for remote analysis, has made NIR spectroscopy an invaluable tool for academic research and industrial quality control. Metrohm Process Analytics offers multiple NIR process analyzers that facilitate non-destructive, accurate measurements, directly in process streams.

In the industrial world, online and inline analysis are the preferred solutions for «real-time» process monitoring without interference. The use of fiber optic probes and flow cells with NIRS systems has opened up new horizons for process monitoring. They make it possible to economically perform remote, unsupervised measurements. This reduces operational cost and increases productivity.

For compact spaces, NIR process analyzers can be placed at a distance from the sampling point with the use of low-dispersion fiber optic cables. Additionally, selecting and implementing the correct optical probes and flow cells is crucial to obtain successful «real-time» results from a process.

With more than 45 years of experience, Metrohm Process Analytics can design the best solutions for your process. If required, up to 10 sampling points can be multiplexed with each analyzer – providing even more information about your process and its quality. Since sample conditions are dependent on the applications, selecting the right sampling solution to use with an NIR system is crucial for successful process implementation.

This brochure is the perfect guideline for choosing the correct sampling solution for the best analytical performance. Identifying and implementing the right fiber and probes has never been easier. Our sales specialists work with you to ensure that the system you select provides the highest quality and most reliable data. The process of selecting the ideal sampling solution is a team effort, requiring the full collaboration of all stakeholders: the user, the integrator (if any), and Metrohm Process Analytics.

The Complete Solution for Your Process

NIRS is a fast, reagent free, non-destructive analysis technique used in many industries for in-process control of manufacturing operations and product quality measurement.


In addition to process analyzers, Metrohm Process Analytics offers complete solution packages for seamless integration: optical fibers, sampling solution (flow cells and/or probes), Industrial PC (IPC), and process communication protocol (digital and/or analog). Implementing online and inline process analyzers requires experience in broad range of technical fields. In the last 45 years, we have developed exactly that.

KEY FEATURES OF METROHM PROCESS ANALYZERS

- Multiplexing capability of NIR systems – Save more on every measurement point and every result.
- No reagents – No chemical waste.
- Wide scan range (1000 to 2250 nm) that enables a wide range of applications.
- Fast analysis results around the clock (24/7) for any sample (liquids, solids, and gases).
- Robust industrial design for operation in harsh and hazardous environments.



SAMPLING POSSIBILITIES FOR REACTION MONITORING WITH PROCESS NIR ANALYZERS



SAMPLING SOLUTIONS

Highly precise measurements do not only depend on the analytical process analyzer. In fact, the type of sampling is just as important to obtaining high quality results. Depending on the application, each sample is subjected to different conditions across various properties, such as temperature and pressure. Furthermore, the chemical composition and physical properties are equally important, since they also influence which type of sampling is necessary.

The key factor in determining a sample is to obtain reproducible results. This ultimately requires a representative sample of the same quality in front of the probe at all times. This can be a challenge. Metrohm Process Analytics has decades of experience in selecting the best sampling solution for a given sample in a specific process. Consult our experts to find the best option for your process.

BENEFITS OF USING PROBES AND FIBERS WITH NIRS PROCESS ANALYZERS

- **Customizable sampling systems** according to the nature of the sample. Solid, liquid, and gas samples can be measured on the same instrument.
- **Best analytical performance** for your process. Each sampling point has unique process conditions.
- **Measure everywhere, close by or from a distance.** Sampling points can be hundreds of meters apart (e.g., two sampling points can be 400 m from each other with the instrument in between).

Versatility of Fiber Optic Probes for Process Monitoring

Chemical analysis in process streams is not always an easy task. The chemical and physical properties, such as viscosity and flammability of the sample streams, can interfere with the analysis measurements. Some industrial processes are quite delicate – even the slightest changes to the process parameters can lead to significant variability in the properties of the final products. It is therefore essential to continuously measure the properties of the stream and adjust the processing parameters via rapid feedback to assure product quality and consistency.

NIR spectroscopy is well suited to this task, providing «real-time» structural and kinetic process data. It is rapid, requires little or no sample preparation and minimal technical expertise to operate. Many of these benefits are due in no small part to its most common method of interfacing – a **fiber optic probe** placed directly into the reactor system. This eliminates the need for sample preconditioning, and consequently reduces time and experimental error associated with it.

ADVANTAGES OF FIBER OPTIC PROBES

The use of fiber optic probes in NIR spectrometer systems has opened up new perspectives for process monitoring. A NIR probe connected to a spectrometer via optical fiber allows direct online and inline monitoring, without interfering with the production process. Fiber optic probes can be placed in very harsh working environments, while the spectrometer and analysis computer remain safe and secure in a shelter. In fact, remote monitoring can be achieved at large distances without significant impact on the signal-to-noise ratios.

Nowadays, a wide variety of NIR optical probes are available. From transmission pair probes and immersion probes, to reflectance probes suitable for contact and non-contact measurements. This diversity allows NIR spectroscopy to be applied to almost any kind of sample composition. This includes melts, solutions, emulsions, suspensions, gas, and powders.



FINDING THE RIGHT MATCH



Selecting the right probe or sample interface to use with a NIR process analyzer is crucial to successful process implementation for inline or online process monitoring. For the system to perform optimally, the probe must be optically matched with the spectrophotometer and with the optical fiber that transmits the spectral data.

Some of the challenges faced by NIR sampling interfaces for process monitoring include sustained high temperatures, sudden changes in temperature, extreme pressures, turbulent flow issues, and fouling of the probes. In order to succeed and overcome these issues, the probe must be selected based on the application, the measurement method, and the environment in which it is to be used. The material of the probe must be inert to the chemicals or materials with which it will be in contact, and yet it must withstand extreme process conditions.



Measurement Principles


NIR spectroscopy can be used to analyze different types of samples. Choosing the right measurement method, sampling module, and accessories is the most important step to developing robust NIR methods.



Transmission

- Sample type: clear liquids, suspensions, and solutions


The sample is placed between the NIR light source and detector. NIR light is transmitted through the sample. The non-absorbed NIR energy continues through to the detector.



Reflectance

- Sample type: flat surfaces, powder, polymer pellets, fluffy materials


The NIR light is directed on the sample which interacts with the light and reflects the non-absorbed light back to the detector. The sample can either be at a distance from a few millimeters to 30 cm from the light source – referred to as a «non-contact measurement». However, when the sample is in contact with the lens, this is referred to as an «inter-actance reflectance measurement».



Diffuse transmission

- Sample type: solid dosage forms without sample preparation (e.g. tablets and capsules)


NIR light penetrates into the sample and interacts with the sample. Due to the particles, the light is scattered throughout the sample. The non-absorbed NIR light is transmitted through the sample reaching the detector.



Diffuse reflectance

- Sample type: solid samples without sample preparation (e.g. creams, pastes, granulates, coarse, fine powders)

NIR light penetrates into the sample and interacts with it. The NIR energy that is not absorbed is reflected back to the detector.



Transflectance

- Sample type: liquids and gels

This measurement method is a combination of transmission and reflectance. A reflector is placed behind the sample to reflect the non-absorbed NIR light back to the detector.

Dedicated Fibers for Optimal Performance

FIBER OPTIC INTERFACING

Generally, NIR light from the instrument is transferred to the process probe using fiber optic cables. As the light scattering properties of the process sample increases, the number of fibers used in the fiber optic bundle must be increased in order to maintain analytical performance. All fibers are made of a low-dispersion material to minimize NIR absorption of the fiber itself. Nevertheless, the fibers absorb light strongly above 2250 nm. Therefore, process NIR instruments with fibers use a maximal wavelength of 2250 nm.

Microbundle fibers are used to monitor slightly scattering liquid media, suspensions, and drying processes.


Microbundles are composed of 37 fibers with a 200 µm diameter silica core for illumination and 7 fibers with 200 µm diameter silica cores for collection. The termination on the instrument consists of SMA 905 connectors and the termination on the sampling side is a metal cladding which is inserted into a probe. The illumination bundle forms a circle when the collection fibers are bundled in the center. Thus the illumination light and the collection light follow the same path.

Single fibers are typically employed to analyze clear liquids. Single fibers have a 600 µm silica core made of single fibers, which are protected with a double layer PVC tubing with Kevlar strain relief. There is a gas-tight gel filling between the layers. Termination on both sides consist of SMA 905 connectors. Single fibers can have lengths up to 200 m. The use of longer fiber optic lengths can enable a process analyzer to be located outside of electrically classified or safety classified areas, and can keep it isolated from harsh operating conditions like large temperature variations.

Metrohm Process Analytics designs the best solutions to provide you with the best results. In the following pages, you will find the optimal solution that we offer depending of the nature of your sample. The sample properties define the sampling solution (probe/flow cell), which in turn defines the type of fibers used.

Liquid Samples

In fast loop, bypass, reactors, and tanks



CLEAR AND TRANSPARENT LIQUIDS

Clear and transparent liquids are easy samples to measure with NIRS, since the light passes easily through the sample without any disturbance. For these types of samples, we recommend using either flow through cells for measurements in transmission, or probes for measurements in, transmittance depending, on the various possibilities within your process.

Samples in fast loop, bypass process pipes

Flow through cells can be installed in a bypass line or a sample preconditioning panel to measure liquid samples. The advantages are clear:

- The flow rate can be controlled to avoid turbulent flow. Laminar flow is preferred for reproducible measurements.
- Bubbles are not an issue, they can be eliminated.
- It's possible to implement a preconditioning panel with filters before analysis. This eliminates particles that might interfere with the measurements.
- Easy access to the flow cell (if cleaning is necessary).
- Samples can be collected at the same point as the NIR system, in order to do model building and/or validation.



Industrial flow cells for harsh industrial environments.

Flow through cells work perfectly with single fibers. The **optical path length** (OPL) is fixed, but it can be adjusted on some flow cell models. For protic solvents (e.g., alcohols, amines), an optical path length of 4 mm is recommended, with 5 mm recommended for aprotic solvents (e.g., alkanes, DMF, DMSO).

For aqueous samples with water content above 95 wt%, an OPL of 0.5 mm is advised. Each time the water content decreases by 10 %, 0.05 mm can be added to the OPL (e.g., for a sample with a maximum of 55 % water, we select an OPL of 0.7 mm).



Industrial flow cell for fast loop and bypass process pipes.

Depending on the requirements, several flow cells are available with different options, such as temperature control or different materials. Flow cells are usually connected to the process with Swagelok®, but flanges can also be used.

Samples in reactors, process pipes and tanks


Some chemical processes consist of multiple reactor vessels, long process pipes and immense tanks in which flow cells are difficult to implement.

In such cases, industrial probes are the right solutions to perform the NIR measurement. There are two types of measurement principles that can be used to measure liquid samples: transmission (the same as in flow cells), or transmittance.



Top: Single fiber transmittance probe with a set of mirrors for defined OPL (left to right): 0.5 mm, 1 mm, 2 mm, 5 mm, 10mm. Bottom: Microbundle fiber transmittance probe with adjustable mirror.

Mirrors with pre-set OPL are available for single fibers, but for microbundle fibers this is not the case. A mirror slides along the probe axis to adjust the OPL. When an application requires high precision measurements, microbundle fibers are the preferred selection. Similar to the flow cells, probes can be connected to the process stream via Swagelok®.



Transmission probe, left: fiber connection; right: sampling with a 2 mm OPL.

Transmission probes work with single fibers and true transmission measurements. The determination of the OPL is identical to that of flow cells, and is fixed for these types of probes. Transmission probes are also called «single pass» probes.

The light passes twice through the sample with **transmittance probes**.¹ Thus the measured gap in the probe is actually half of the OPL (e.g., a 1 mm transmittance probe corresponds to a 2 mm OPL). The liquid sample is contained between a lens and a concave gold-coated mirror. Therefore, the light is focused on the sample, and the signal-to-noise ratio is higher than for transmission measurements (and thus more precise).



Example of process connections. Left: Transmission probe with welded flange. Right: Transmittance probe with welded flange.

¹ Note that for the same sample and OPL, the spectrum from a transmittance probe will have 25 % more absorbance than the spectrum measured in transmission. Transmittance probes work with single fibers or microbundle fibers.

Liquid Samples

In fast loop, bypass, reactors, and tanks



CLEAR AND TRANSPARENT LIQUIDS – HARSH CHEMICALS

Sampling solutions made of metals (e.g., stainless steel) are not always compatible with all applications. In cases where inorganic acids, such as sulfuric acid (H_2SO_4) or hydrofluoric acid (HF), are present in the sample stream, other materials are required for the sampling solutions. Perfluorinated polymers are compatible with such substances. Here we describe some sampling solutions made of polytetrafluoroethylene (PTFE). Be aware that other types of perfluorinated polymers can also be used (e.g., perfluoroalkoxy alkanes, PFA; or polyvinylidene fluoride, PVDF).

Single fiber flow cell (adjustable OPL with spacers)

Transmission measurements can be performed with a flow-through cell made of PTFE. The same rules apply for the determination of the OPL as if it was made of stainless steel 316L. This PTFE configuration allows sample contact with a sapphire window, which is inert to most chemicals. Optionally, an extra thin layer of PTFE can be added to the window so that the sample is only in contact with PTFE. For applications with high water content, a path length between 0.5 and 1.0 mm is preferred.



PTFE single fiber flow cell on top of a stand.

Transparent process pipes (PTFE Clamp-on flow cell)

When modifications on the process line are not possible (e.g., using a bypass), or the sample stream cannot be in contact with the sampling solution (e.g., flow cell), a PTFE clamp-on flow cell is the right solution. This clamp-on flow cell works with single fibers and is simple to implement. Just clamp it on to a pre-existing tubing at your production site. The OPL depends on the size of the sample stream tubing.

Result quality is non-proportional to the diameter of the process tubing. Rather, as the tube decreases in size, its curvature increases. This interferes with the results.



PTFE single fiber clamp-on flow cell.

PTFE transreflectance probe

Probes are also in contact with the sample stream, so it is important to consider their chemical compatibility. These probes work perfectly with microbundle fibers, and the OPL is adjustable by moving the mirror position.

CLEAR AND TRANSPARENT LIQUIDS – DEMANDING PROCESSES

Transmission probe pair

Some applications are quite demanding and require specific solutions for reliable analysis. Sample streams at elevated temperatures and pressures are particularly challenging for regular flow cells. An alternative solution is transmission probe pairs from Metrohm Process Analytics, which can withstand up to 250 °C and 5000 psi. Transmission probe pairs work with single fibers or microbundle fibers, the latter giving about 20% more signal for the same sample and OPL.

In addition, transmission probe pairs are also available with threads, making them especially suitable for polymer extrusion applications or for applications requiring up to 30,000 psi.

High temperature probes

For sample temperatures above 250 °C, the probe design includes an inlet for cold compressed air to cool the tip of the internal fiber and increase the fiber lifetime. Transmission and transreflectance microbundle probes are suitable for high temperature processes.

Although it sounds complicated, implementing these probes is extremely simple and straightforward. Transmission probe pairs can be directly inserted in a cross section of a process pipe, connected through a Flange, a Swagelok® or a combination of both. For installation on smaller process pipes, flow cells that allow for probe insertion are available. (Same situations applies for transreflectance probes).

PTFE PROPERTIES:


- Hydrophobic
- Chemically inert
- Nonreactive to almost all known chemicals
- Excellent thermal resistance
- Nonadhesive



Top: Transmission probe pairs with spacer. Top center: single fiber. Bottom center: Transmission pair with thread for implementation in extruder.

Low temperature probes

For temperatures from 0 °C down to -110 °C, the probe design includes an inlet for dry nitrogen to prevent condensation and the formation of ice crystals inside the probe. Special O-rings adapted for such temperatures are included as well. Single fiber single pass transmission probes can be used for low temperature processes.



Top left: flow cell for transreflectance probes. Top right: flow cell for transmission probe pairs. Bottom: example of a transmission probe pair implementation directly on a process pipe.

Solid Samples

Powders, crystals, tablets and more



FINE POWDERS

Reflectance probes for fine powders

For measurements of fine powders, the reflective power of the sample itself is used to our advantage. The particles can act as small mirrors. Thus inter-actance reflectance probes are the best solution to measure this type of sample. The NIR light absorbed by the solid particles constitutes the spectra obtained by the process analyzer. Additionally, the scattered light is measured, influencing the position of the spectrum baseline. It is therefore possible to measure particle size by analyzing the baseline position on the spectra. Other physical parameters of solid particles, such as melt index or intrinsic viscosity, can also be measured. The bandwidth of the peaks is influenced by those physical parameters and can be quantified as well.

The **reflectance probe** has a sapphire lens that is directly in contact with the metal (no epoxy or O-ring in contact with the sample). This probe can be used in conditions up to 250 °C and 5000 psi and with additional options can withstand up to 400 °C. The lens focuses the light on the surface of the sapphire. Therefore, the sample needs to be in contact with the sapphire to obtain precise and reproducible spectra.

Sample color is also important, especially when it comes to carbon black powders. Black powders influence the amount of reflected light and very often these samples cannot be measured. However, it can happen that the powder is black but composed of metallic particles, in which case they will act as actual mirrors.



Example of reflectance probes, from top to bottom: reinforced PTFE spoon probe, regular probe, 45° angle probe, SS spoon probe, and PTFE spoon probe.

The window of the probe has a small diameter. Some tricks can be used to ensure the best contact:

- When the sample stream is vertical (e.g., the sample is in a dryer or falling down a process pipe), spoon probes are the best solution. The measurement is performed when a sufficient amount of sample falls in the spoon. After the measurement is complete, compressed air is used to remove the sample so that a fresh sample can be collected. This spoon probe is available in stainless steel 316 L or alternatively in PTFE. The contact surface with the sample is exclusively PTFE. A small layer of PTFE covers the lens, and the light passes through this layer.

This is particularly useful if the sample has etching properties or if glass, lenses, or windows are prohibited in a process.

- 45° probes are used for powder flowing in a horizontal process pipe or in a blender with a mechanical stirrer. Since the sample can be flowing in different directions, the 45° angle on the probe nozzle ensures good contact between the sample and the lens.




Illustration of different type of reflectance probes.



Spoon probe implemented in a dryer.

OPAQUE SAMPLES

Diffuse reflectance and diffuse transmission measurements for particles

Some samples such as particles or polymer films are opaque but the light can still pass. They appear opaque, milky or white due to the scattering of light within the sample. For these types of samples, a diffuse type of measurement is used. We recommend using the microbundle transmission probe pair for polymer films and opaque samples and diffuse reflectance probe for particles and porous samples.



Diffuse reflectance probe.

Solid Samples

Pellets, crystals, ores, and more



MACROSCOPIC PARTICLES

Non-contact probes for large particles

Because of their bigger shape and non-uniformity, it is almost impossible to use contact probes to measure macroscopic solid particles such as polymer pellets, wood chips, and crystals, among others.

For such samples, a non-contact probe is preferable for inline analysis. This probe is unique, as the fibers are permanently attached to it. Microbundles consist of 6 x 15 fibers for illumination, and 1 x 91 fibers for collection.

The illumination fibers form a focal plane approximately 8 mm away from the surface of the probe. All particles between the focal plane and the probe will act as a mirror and reflect the light back to the detector.



Non-contact probe used for the analysis of large non-uniform particles.



For very soft samples such as wool, a spoon sleeve can be used to collect the sample and then remove it with compressed air, same as with fine powders.



Example of non-contact probe inserted in an outer sleeve for fluffy samples.

Even if the principle is called non-contact, the probe is still in contact with the sample. It can be used in non-contact mode through a window as long as the sample is in permanent contact with the window.

- Light beam
- Lens
- Sapphire window
- Illumination focal plane
- Detector



Sample in the illumination cone will reflect light towards the detector.

Illustration of the non-contact probe measurement principle.

Suspensions and Slurries – Solid or Liquid?

SUSPENSION MEASUREMENTS

> 30 wt% solid content

Sample streams with a relatively large amount of solids can be a challenging to quantify. The measurements have to be done in reflectance mode and the solid part can be measured easily. For the liquid part of the sample stream, measurement depends on its properties. If the sample is water-based, the solid particles need to be removed before analysis. If the sample is non-aqueous, the liquid content can be quantified concurrently with the solid content.



< 30 wt% solid content

Sample streams with low solid content can be measured using a transmission probe. Since the solid affects only the baseline and peak bandwidth, the solid part can be quantified along with the liquid solution. Aqueous solutions can be challenging due the small OPL used and the risk of probe fouling.



Metallic particles

Sample streams that contain metallic particles are good reflectors and are therefore easy to quantify. However, because of their reflective power, the flow rate has a large influence on the NIR spectra. For the most reproducible measurements, flow rate should be kept constant.



Gas Samples

Continuous and batch-wise analysis



There are different ways to treat gas samples for analysis. If they can be liquefied, they are treated as liquid samples. If they remain gaseous, then there are two approaches which can be used depending on the process setup:

- For **continuous** analysis, the gas sample stream is run through a flow cell at a constant flow rate and temperature. Like clear transparent liquids, the measurement is carried out in transmission mode so that absolute concentration values can be obtained.
- Batch-wise** analysis is also an option to measure gas streams. First, the exit valve of the flow cell is closed, so the sample pressure builds up within the flow cell. When the pre-defined pressure is reached, the valve at the entry of the flow cell is closed as well so that analysis can be resumed. The measurement is made in transmission mode at constant pressure and temperature, thus relative concentrations are obtained. This strategy is used to measure low concentrations.

Molecular density is an important parameter in gas samples. The density of the molecules in a gas sample is lower than in liquid samples. Therefore, longer OPL are used that range from a few centimeters to meters. The OPL depends on the required precision and concentration levels.

Near-infrared spectroscopy is a secondary technique, meaning it requires a prediction model (obtained by a primary/reference method such as titration or HLPC) when quantifying samples. For gas samples, gas chromatography is the preferred reference method. A single fiber transmission probe pair or a microbundle transmission probe pair can be used for such gas applications. The microbundle version gives better accuracy due to a higher signal-to-noise ratio.

It is important to note that gas samples follow the ideal gas law: the concentration is directly proportional to the **pressure** and inversely proportional to the **temperature**. Therefore, these two parameters need to be precisely controlled for accurate and safe analysis.




Illustration of a gas flow cell implementation.

Selection of Sampling Solutions

Because every production plant, process, and sample is unique, there is no universal rule for the selection of the ideal sampling solution. The sampling solution is as important as the technology itself. Appropriate sampling will result in good reproducibility and precision. Here are some guidelines on how to **select** the best sampling solution:

1

MEASUREMENT

First, consider the phase of the sample that determines which measurement principle to select:

- Transmission (Transflectance) for liquids, gasses, and some suspensions.
- (Diffuse) Reflectance for powders and some suspensions.
- Non-contact measurement for plain solids and coarse grains.

MEASUREMENT PRINCIPLE REFINING

Second, consider the constituents to be measured and the precision needed. These parameters give information about the optical path length to select for transmission/transflectance measurements or about the measurement distance for non-contact measurements.

2

3

ROBUSTNESS/RESISTANCE

After the measurement principle is defined, the type of probe or flow cell must be selected. The sample properties (corrosive, abrasive) and conditions (temperature/pressure/viscosity/flow rate) need to be considered to select the correct sampling solution and construction material. Standard metals used include SS316, Hastelloy C, Titanium 6, Monel, and nickel.

FIBERS

The type of fiber to be used will be deduced from the steps above. It is important to consider protection for the fibers with a PVC conduit, for instance, to prevent mechanical and/or chemical damage. Usually, microbundle fibers are used for higher signal-to-noise and precision.

4

Implementation of Sampling Solutions

The placement of a probe or a flow cell is not trivial. Very often, the existing industrial process has limitations such as space constraints or safety regulations to comply to. Here are some guidelines on how to **implement** sampling solutions:

PROBES

Probes that measure in-situ are connected directly to the process via flanges and/or Swagelok®. Such immersion probes are permanently installed. This makes them more difficult to maintain. However, they are typically cleaned by circulating or pulverizing a cleaning solution on the probe surface or with the use of steam. Care must be exercised when using steam due to its extreme temperature. When the only option for sufficient probe cleaning is the probe from the process itself, implementing a flow cell is much more practical. In doing so, the flow cell is isolated from the main process stream avoiding any disturbance to normal process operations while cleaning the probe. Solutions for probe extraction are available if needed.

The immersion depth of the probes depends on the type of flow within the process. The ideal immersion depth should be where the sample is the most representative and therefore process dependent.

FLOW CELL

Flow cells are usually installed on fast loops or on bypass streams. Differences in flow rates and process pipe size have to be considered to avoid turbulent flows or the presence of bubbles. Flow cells are preferably installed in a box isolated from external conditions. Flow cells offer the possibility to use preconditioning panels prior to the flow cell to prepare the sample (e.g., filtration, extraction) and to achieve good sample reproducibility. Another advantage to integrate flow cells within sampling panels is that validation can be easily implemented in two different ways: known sample concentrations can be injected directly, while manual samples can be collected at the same point for laboratory measurements. Temperature control is also easier with flow cells. It is recommended to have the sample flowing upward to avoid gravity effects or settling effects when measuring suspensions.



Timeline for Sampling Solution Deliveries

Some probes are standard items. But, in more than 90% of the cases, the sampling solution is customized to the process requirements. The selection and ordering of the sampling solution is a multi-step process that involves many parties. This process should not be overlooked, but rather addressed from the beginning of a project.

SAMPLING:

- Laminar flow – reproducible measurements.
- Filtration options (particulate/ water) to avoid interferences.
- Temperature control for better accuracy.
- Easy accessibility to probe (for cleaning or running calibration standards).
- Ability to collect the sample at the same point where the NIR instrument collects spectra (for calibration, validation, or updating of models).

