

## 1 Declaration of Conformity

<i>Description</i>	Comprehensive and user-friendly PC program for the determination of dispensing accuracy and dosing errors of Metrohm Dosing an Exchange units according to ISO 8655 standard. Dialog in English and German.
<i>Quality system</i>	The software mentioned above was developed in accordance with the requirements of the ISO 9001 quality system regarding the design, testing and servicing of <b>Metrohm software</b> .
<i>Technical specifications</i>	The technical specifications are documented in the online help.
<i>Software validation</i>	The software was validated with respect to functionality, analytical performance and accuracy of results. The software functions are documented in the online help.

**Herisau, April 1, 2022**

---

P. Hunziker  
Vice President,  
Head of Development

Dr. M. Saba  
Vice President,  
Head of Quality Management &  
Regulatory Affairs