

1 EU declaration of conformity / UK declaration of conformity

This declaration attests the compliance of the listed products with the standard specifications for electrical instruments and accessories.

The responsibility for issuing this declaration of conformity lies solely with the manufacturer.

1.1 Product validity

This declaration is valid for the following products or product versions:

Article number	Designation
6.0502.100	Ion-selective electrode, Br
6.0502.120	Ion-selective electrode, Cl
6.0502.130	Ion-selective electrode, CN
6.0502.140	Ion-selective electrode, Cu
6.0502.150	Ion-selective electrode, F
6.0502.160	Ion-selective electrode, I
6.0502.170	Ion-selective electrode, Pb
6.0502.180	Ion-selective electrode, Ag/S

1.3 **Electromagnetic compatibility (EMC)**

Design Engineering and type testing

- **EN 61326-1: 2021**
Electrical equipment for measurement, control and laboratory use - General EMC requirements

1.3.1 **Emission**

Standards fulfilled

- EN 55011 / CISPR 11: 2020

1.3.2 **Immunity**

Standards fulfilled

- EN 61000-4-3: 2020

1.4 **Manufacturer**

Metrohm AG, Ionenstrasse, CH-9100 Herisau/Switzerland

Metrohm AG holds the SQS certificate ISO 9001: 2015 for quality management systems with scope for the development, production, sale and service of instruments and accessories for titration, ion chromatography, spectroscopy and electrochemistry, including software solutions.

Herisau, March 19, 2026



Patrick Hunziker
Vice President,
Head of Development



Dr. Miriam Saba
Vice President,
Head of Quality Management
& Regulatory Affairs