

Project 20816800
Audit Report: 21 CFR Part 11



Metrohm AG
CH-9101 Herisau
Switzerland
Tel. +41 71 353 85 85
Fax +41 71 353 89 01
www.metrohm.com

***tiamo 2.0* for titration**

Audit Date:	13.01.2009
Auditor:	Sieghard Wagner, mech. Engineer (grad.), Chemengineering Business Design GmbH
Audit Target:	Compliance check of <i>tiamo 2.0</i> software for controlling titration devices against the requirements of 21 CFR Part 11.
Description:	<p>PC software for controlling titration devices. Creation of methods, carrying out determinations (titrations and measurements), data acquisition, evaluation and archiving in a database.</p> <p>The above software was developed and manufactured in accordance with the requirements demanded by the ISO 9001 quality system regarding the design, manufacture and servicing of Metrohm instruments.</p>
Investigation:	<p>The functions and properties of the above software system are audited with the requirements of 21 CFR Part 11 and the current interpretations.</p> <p>During the audit the following items were examined: validation, documentation, audit trail, electronic copies, access rights, user access, password, access violation, sequence of steps, plausibility, device check, data encryption, digital signature.</p>
Operator Responsibility:	<p>The compliance with the requirements of 21 CFR Part 11 can only be met together with the operational environment. The requirements for such an environment are:</p> <ul style="list-style-type: none">• technical environment, data management• training• administration• SOP (Standard Operating Procedures)
Summary:	<p>The software is compliant to the following 21 CFR Part 11 requirements:</p> <ul style="list-style-type: none">• 11.10.(a), (b), (d), (e), (f), (g), (h)• 11.30, 11.50, 11.70• 11.100 (a), 11.200 (a), 11.300 (a), (b), (d) <p>The software is compliant to the following 21 CFR Part 11 requirements with support of the operator.</p> <ul style="list-style-type: none">• 11.10 (c), (i), (j), (k)• 11.100 (b), 11.300 (c), (e) <p>The software is supporting digital signature.</p>