

1 Declaration of Conformity

Description

Comprehensive and user-friendly PC program for the freely programmable control of ion chromatography networks comprising instruments of the Professional IC line and other devices. Creation of methods, carrying out determinations, storage and recalculation of results. Completely compliant with FDA 21 CFR Part 11. Client/server operation possible. Dialogs in English and German.

- **6.6059.310 MagIC Net 3.1 Basic**
Limited functionality, supports only the 883 Basic IC and the 863 Compact Autosampler; no client/server support.
- **6.6059.311 MagIC Net 3.1 Compact**
Limited functionality and limited device support; no client/server support.
- **6.6059.312 MagIC Net 3.1 Professional**
Full functionality and device support; no client/server support.
- **6.6059.313 MagIC Net 3.1 Multi**
Full functionality and device support; includes client/server support.

Quality system

The software mentioned above was developed in accordance with the requirements of the ISO 9001 quality system regarding the design, testing and servicing of Metrohm software.

Technical specifications

The technical specifications are documented in the installation manual.

Software validation

The software was validated with respect to functionality, analytical performance and accuracy of results. The software functions are documented in the online help.

Herisau, September 1, 2014



P. Hunziker
Vice President,
Head of Development



U. Kürsteiner
Head of Quality Management