

MagIC Net 4.0



Metrohm AG
CH-9100 Herisau
Switzerland
Tel. +41 71 353 85 85
Fax +41 71 353 89 01
<https://www.metrohm.com/en>

Audit Date:	07-June-2021 ¹
Auditor:	Sieghard Wagner, mech. Engineer (grad.), Chemengineering Germany GmbH
Audit Target:	Compliance check of <i>MagIC Net 4.0</i> software for controlling ion chromatography devices against the requirements of 21 CFR Part 11.
Description:	PC software for controlling ion chromatography devices. Creation of methods, performance of analyses (chromatograms and measurements), data acquisition, evaluation, and archiving in a database. The software was developed by Metrohm AG in accordance with the ISO 9001 requirements regarding design, manufacturing, and maintenance.
Investigation:	The functions and properties of the above software system are audited with the requirements of 21 CFR Part 11 and the current interpretations. During the audit the following items were examined: validation, documentation, audit trail, electronic copies, access rights, user access, password, access violation, sequence of steps, plausibility, device check, data encryption, electronic signature.
Operator Responsibility:	The compliance with the requirements of 21 CFR Part 11 can only be met together with the operational environment. The requirements for such an environment are: - technical environment, data management - training - administration - Standard Operating Procedures (SOP).
Summary:	The software is compliant to the following 21 CFR Part 11 requirements: 11.10 (b), (d), (e), (f), (g) 11.50, 11.70 11.300 (a) The software is compliant to the following 21 CFR Part 11 requirements with support of the operator: 11.10 (a), (c), (h), (i), (j), (k) 11.100 (a), (b), 11.200 (a), 11.300 (b), (c), (d) The software supports electronic signatures.



Sieghard Wagner

¹ This 21 CFR Part 11 certificate is based on an on-site audit performed January the 13th 2009. Subject of this audit was the platform commonly used by MagIC Net 2.x and tiamo 2.x. This platform provides the functionality for electronic record and signature processing. According to Metrohm AG, the implemented changes in the current version are not relevant with regard to 21 CFR Part 11 or compliant with 21 CFR Part 11 (see Release Notes 8.102.8016EN, 8.102.8024EN, 8.102.8033EN, 8.102.8039EN, 8.102.8048EN, 8.102.8060EN, 8.102.8072EN, 8.102.8082EN, 81028090EN). Therefore, this update does not require a physical re-audit.