

Declaration of Conformity



Metrohm Ltd.
CH-9101 Herisau
Switzerland
Tel. +41 71 353 85 85
Fax +41 71 353 89 01
www.metrohm.com

for the instrument

828 IC Dual Suppressor

Description

Combined electrochemically regenerated «packed bed» and CO₂ suppressor. High capacity, designed for use in gradient systems.

The above instrument was developed and manufactured in accordance with the requirements of the ISO 9001 quality system regarding the design, manufacture and servicing of Metrohm instruments.

The instrument was manufactured and tested according to the following standards:

Electromagnetic compatibility: emission

IEC 61326-1, EN 55022, CISPR 22

Electromagnetic compatibility: immunity

IEC 61326-1, IEC 61000-4-2, IEC 61000-4-3, IEC 61000-4-4, IEC 61000-4-5,
IEC 61000-4-6, IEC 61000-4-11, IEC 61000-4-14, Namur

Safety specifications

IEC 61010-1, UL 3101-1

The instrument has also been certified by the Swiss Electrotechnical Association (SEV), which is member of the International Certification Body (CB/IEC).

The instrument meets the requirements of the CE mark as contained in the EU directives 89/336/EEC and 73/23/EEC and fulfils the following specifications:

EN 61326-1	Electrical equipment for measurement, control and laboratory use – EMC requirements
EN 61010-1	Safety requirements for electrical equipment for measurement, control and laboratory use

The technical specifications are documented in the instruction manual.

The instrument was validated with respect to functionality, analytical performance and accuracy of results. The instrument functions are documented in the instruction manual.

Herisau, 20 March 2002

Dr. J. Frank
Development Manager

Ch. Buchmann
Production and
Quality Assurance Manager

Quality Management Principles

Metrohm Ltd. holds the ISO 9001 Certificate, registration number 10872-02, issued by SQS (Swiss Association for Quality and Management Systems). Internal and external audits are carried out periodically to assure that the standards defined by Metrohm's QM Manual are maintained.

The steps involved in the design, manufacture and servicing of instruments are fully documented and the resulting reports are archived for ten years. The development of software for PCs and instruments is also duly documented and the documents and source codes are archived. Both remain the possession of Metrohm. A non-disclosure agreement may be asked to be provided by those requiring access to them.

The implementation of the ISO 9001 quality system is described in Metrohm's QM Manual, which comprises detailed instructions on the following fields of activity:

Instrument development

The organisation of the instrument design, its planning and the intermediate controls are fully documented and traceable. Laboratory testing accompanies all phases of instrument development.

Software development

Software development occurs in terms of the software life cycle. Tests are performed to detect programming errors and to assess the program's functionality in a laboratory environment.

Components

All components used in the Metrohm instruments have to satisfy the quality standards that are defined and implemented for our products. Suppliers of components are audited by Metrohm as the need arises.

Manufacture

The measures put into practice in the production of our instruments guarantee a constant quality standard. Production planning and manufacturing procedures, maintenance of production means and testing of components, intermediate and finished products are prescribed.

Customer support and service

Customer support involves all phases of instrument acquisition and use by the customer, i.e. consulting to define the adequate equipment for the analytical problem at hand, delivery of the equipment, user manuals, training, after-sales service and processing of customer complaints. The Metrohm service organisation is equipped to support customers in implementing standards such as GLP, GMP, ISO 900X, in performing Operational Qualification and Performance Verification of the system components or in carrying out the System Validation for the quantitative determination of a substance in a given matrix.