

Audit Report: GAMP5® Software Categorisation

900 Touch Control

5.900.0041



Metrohm AG

CH-9100 Herisau


Switzerland

Tel. +41 71 353 85 85

Fax +41 71 353 89 01

<https://www.metrohm.com/en>

- Date:** 2019-02-08
- Author:** Sieghard Wagner, mech. Engineer (grad.), Chemengineering Business Design GmbH
- Objective:** Classification of *900 Touch Control* (version 5.900.0041) according to the GAMP5® software categories.
- Description:** *900 Touch Control* is an operating unit for the Titrandos, USB Sample Processors, 856 Conductivity Modules, 867 pH Modules and 846 Dosing Interfaces with touch-sensitive display for processing methods, analysis (titration, measurements), data acquisition, evaluation, and reporting.
900 Touch Control was developed by the Metrohm AG in accordance with ISO 9001 requirements regarding design, manufacturing, and maintenance.
- Categorisation:** The *900 Touch Control* firmware is a “Non-Configured Product” – as such, it is categorized into **GAMP software category 3**.
- Justification:
The firmware configuration is limited to:
- Customization of the system’s runtime environment, e.g.:
 - Maintenance of master data (methods, sample data, etc.)
 - Setup of technical parameters (connected devices, etc.)
 - Configuration of security settings
 - Definition of users and user groups (with pre-defined privileges).
- However, these are no structural modifications or customizations to adapt the firmware to customer-specific business processes.¹
- Creation of methods:
The creation and modification of methods is based on built-in standard system functionality. During normal system operation, methods are adapted to specific analytical procedures on a case-by-case basis. This has to include appropriate checks and verifications, especially of all calculations, settings, and reports included – if applicable. These measures are to be implemented as part of the operational controls in order to maintain the validated state.


Sieghard Wagner

¹ see definition of Software Category 3: GAMP 5, Appendix M4