

1 EU declaration of conformity

This declaration attests the compliance of the instrument with the standard specifications for electrical instruments and accessories and with the standard specifications for safety and system validation of the manufacturing company.

The responsibility for issuing this EU declaration of conformity lies solely with the manufacturer.

1.1 Product validity

This declaration is valid for the following products or product versions:

- **915 KF Ti-Touch**

Compact Titrator for volumetric Karl Fischer titrations with touch-sensitive color display, integrated measuring and stirrer interfaces as well as interfaces for peripheral devices such as dosing devices, additional stirrers and titration stands etc.

1.2 Directives

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:



- **2014/35/EU –Low Voltage Directive, LVD**

Directive 2014/35/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of electrical equipment designed for use within certain voltage limits; Official Journal of the EU L96, 29 March 2014, p. 357-374

- **2014/30/EU – EMC Directive, EMC**

Directive 2014/30/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to electromagnetic compatibility; Official Journal of the EU L96, 29 March 2014, p. 79-106



- **2011/65/EU – Directive for certain hazardous substances, RoHS**

Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment; Official Journal of the EU L174, 1 July 2011, p. 88-110

- **2012/19/EU – Disposal and recycling of electrical and electronic equipment, WEEE**

Directive 2012/19/EU of the European Parliament and of the Council of 4 July 2012 on waste electrical and electronic equipment (WEEE); Official Journal of the EU L197, 24 July 2012, p. 38-71

1.3 Safety specifications

This instrument fulfills the following safety requirements:

Design and type testing

- EN 61010-1: 2010

Safety requirements for electrical equipment for measurement, control, and laboratory use

- **EN 60529: 2013 – degree of protection IP40**

Degrees of protection provided by enclosures (IP Code)

- **ISO 12100: 2010**

General principles for design - Risk assessment and risk reduction

- EN 61140: 2016 – protection class III

Protection against electric shock - Common aspects for installation and equipment

1.4 Electromagnetic compatibility (EMC)

Design and type testing

- EN 61326-1: 2013

Electrical equipment for measurement, control and laboratory use -
General EMC requirements

1.4.1 Emission

Standards fulfilled

- EN 61000-6-3: 2011
- EN 55011 / CISPR 11: 2016

1.4.2 Immunity

Standards fulfilled

- EN 61000-6-2: 2005
- EN 61000-4-2: 2009
- EN 61000-4-3: 2010
- EN 61000-4-4: 2012
- EN 61000-4-5: 2014
- EN 61000-4-6: 2014
- EN 61000-4-8: 2010
- EN 61000-4-11: 2004



- ## 1.5 Manufacturer

Herisau, Oktober 30th, 2018

Head of Quality Management
& Regulatory Affairs

The product is listed in the register of authorizations at ESTI.

 ETL US LISTED
Intertek

■■■■■■■ 3

2.3 IEC certificate



This instrument has been tested in accordance with the IEC standards and certified to the IECEE CB Scheme. The tests also comprise the national differences for Europe, the USA and Canada. Therefore, the instrument is in compliance with the respective standards EN 61010-1, UL 61010-1 and CSA-C22.2 No. 61010-1.

The CB certificate is available in the directory of certified products at electrosuisse.ch.