

Audit Report: GAMP5® Software Categorisation

915 KF Ti-Touch



Metrohm AG
CH-9100 Herisau
Switzerland
Tel. +41 71 353 85 85
Fax +41 71 353 89 01

www.metrohm.com

- Date:** 08-May-2017
- Author:** Sieghard Wagner, mech. Engineer (grad.), Chemengineering Business Design GmbH
- Objective:** Classification of *915 KF Ti-Touch* (version 5.915.0040) according to the GAMP5® software categories.
- Description:** *915 KF Ti-Touch* is a stand-alone system with touch-sensitive display for processing methods, analysis (titration, measurements), data acquisition, evaluation and reporting.
915 KF Ti-Touch was developed by the Metrohm AG in accordance with ISO 9001 requirements regarding design, manufacturing and maintenance.
- Categorisation:** The *915 KF Ti-Touch* firmware is a “Non-Configured Product” – as such, it is categorized into **GAMP software category 3**.
- Justification:
The firmware configuration is limited to:
- Customization of the system’s runtime environment, e. g.:
 - Maintenance of master data (methods, sample data, etc.)
 - Setup of technical parameters (connected devices, etc.)
 - Configuration of security settings
 - Definition of users and user groups (with pre-defined privileges).
- However, these are no structural modifications or customizations to adapt the firmware to customer-specific business processes.¹
- Creation of methods:
The creation and modification of methods is based on built-in standard system functionality. During normal system operation, methods are adapted to specific analytical procedures on a case by case basis. This has to include appropriate checks and verifications, especially of all calculations, settings and reports included – if applicable. These measures are to be implemented as part of the operational controls in order to maintain the validated state.


Sieghard Wagner

¹ see definition of Software Category 3: GAMP 5, Appendix M4