

Manufacturer: FOSS Analytical A/S
Foss Allé 1
DK-3400 Hillerød
Denmark

Instrument: XDS Vial Heater 60072973 91763055
Name Part No. From Serial No.'s

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Has been produced in compliance with:

ElectroMagnetic Compatibility Directive 2014/30/EU

IEC/EN 61000-6-1:2007 Generic immunity standard. Residential, commercial and light industry
IEC/EN 61000-6-3:2007 Generic emission standard. Residential, commercial and light industry
EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use.
Immunity: Basic immunity
Emission: Class B equipment

Low Voltage Directive 2014/35/EU

EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use
EN 61010-2-010:2014 Part 2-010 Particular requirements for laboratory equipment for the heating of materials.

RoHS directive 2011/65/EU

2017-08-08

Date



Per Sonne Sørensen
Senior Vice President Global Quality

Dedicated Analytical Solutions

FOSS Analytical A/S Tel +45 7010 3370
Foss Allé 1 Fax +45 7010 3371
DK-3400 Hillerød E-mail info@foss.dk
Denmark Web www.foss.dk

FOSS Analytical Co., Ltd. Tel +86 512 6292 0100
6 Louyang Road, Building 1 Fax +86 512 6280 5630
215121, SIP, Suzhou E-mail info@foss.dk
P.R. China Web www.foss.dk