

1 Declaration of Conformity

<i>Description</i>	<p>Comprehensive and user-friendly PC program for the freely programmable control of ion chromatography networks. Creation of methods, carrying out determinations, storage and recalculation of results. Dialogs in English, German, Chinese and Traditional Chinese.</p> <ul style="list-style-type: none">▪ MagIC Net 4.1 Basic Limited functionality, supports only the Eco IC and the 863 Compact Auto-sampler; no client/server support, no compliance with FDA 21 CFR Part 11.
<i>Quality system</i>	<p>The software mentioned above was developed in accordance with the requirements of the ISO 9001 quality system regarding the design, testing and servicing of Metrohm software.</p>
<i>Technical specifications</i>	<p>The technical specifications are documented in the installation manual.</p>
<i>Software validation</i>	<p>The software was validated with respect to functionality, analytical performance and accuracy of results. The software functions are documented in the online help.</p>

Herisau, September 27, 2022



Patrick Hunziker
Vice President,
Head of Development



Dr. Miriam Saba
Vice President,
Head of Quality Management &
Regulatory Affairs