



946 Portable VA Analyzer

General information

Software version: 1.0

Instrument:

Sensor: scTRACE Gold

User name:

Report: No

Report elements: -

Method

General

Method name: AB429 Cleaning scTRACE Gold.detp

Remarks: 15 mL H₂O + 3 mL Electrolyte (c(KCl) = 0.3 mol/L, c(HCl) = 0.1)

Determination

Sample volume (mL): 18.0

Total cell volume (mL): 18.0

Stirring time (s): 10.0

Stirring rate (1/min): 2000

Measure blank: No

No. of blanks: 0

Blank value correction: No

No. of replications: 0

No. of additions: 0

Voltammetric

Measuring mode: Linear sweep

Current measuring range: Auto

Cyclovoltammetric pretreatment

Start potential (V): -1.0

Vertex potential (V): 0.8

Potential step (V): 0.01

Sweep rate (V/s): 1.0

No. of cycles: 5

Potentiostatic pretreatment

Potential 1 (V): -0.8

Waiting time 1 (s): 10.0

Potential 2 (V): -0.3

Waiting time 2 (s): 5.0

Equilibration time (s): 5.0

Sweep

Start potential (V): -0.2

End potential (V): 0.6

Potential step (V): 0.01



946 Portable VA Analyzer

Sweep rate (V/s): 0.1

Evaluation

Data processing

Smoothing: 1
Calibration method: Standard addition

Peak evaluation

	-
Characteristic potential (V)	0.0
Tolerance (V)	0.05
Min. width (V)	0.05
Max. width (V)	0.5
Min. measured quantity (µA)	0.001
Baseline type	Linear
Base point automatic	Yes
Start base point (V)	0.0
End base point (V)	0.0

Standard solutions

	-	Volume (mL)
Standard 1	1.0 mg/L	0.1
Standard 2	-	-
Standard 3	-	-
Standard 4	-	-