



Ecotrode

6.0232.100

Combined pH electrode for aqueous acid/base titrations.

Reference electrolyte: $c(\text{KCl}) = 3 \text{ mol/L}$, storage in storage solution.

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 6.0232.100

Order Number	Description	Quantity
6.1236.050	Sleeve with SGJ 14/12 mm Sleeve with SGJ 14/12 mm, polyethylene.	1 Stk
		
6.2008.040	Storage vessel Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units.	1 Stk
		

Optional accessories

6.2308.020

Electrolyte 3 mol/L KCl (250 mL)

Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$, (for Ag/AgCl reference systems)



6.2323.000

Storage solution

Storage solution for all combined pH glass electrodes with reference electrolyte $c(\text{KCl}) = 3 \text{ mol/L}$



6.2325.000

pHit kit

Maintenance kit for pH electrodes

The kit contains:

- 50 mL cleaning solution
- 50 mL 3M KCl solution
- 50 mL storage solution
- 2 Storage vessels
- Instructions for use



6.2325.100

Cleaning solution 3 x 50 mL



Reliable measuring results over long periods of time can be guaranteed only if the pH glass membrane and the diaphragm receive preventive and regular care. Cleaning by etching with toxic chemicals or applying mechanical treatment to the diaphragm is not only complicated and expensive, it also accelerates the ageing of the pH glass electrode.

The cleaning solution was developed for easy and gentle cleaning of pH glass electrodes. Regular use can considerably prolong their service life.

This cleaning solution is also part of the pHit kit.

6.2104.020

Electrode cable / 1 m / F



For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).