

# Combined Au ring electrode

6.0452.100

Combined gold ring electrode with a ceramic pin diaphragm.
This is well suited for redox titrations when the pH value varies, e.g.:

– for titrations using $Hg(NO_3)_2$
– for titrations in presence of chromium or iron

c(KCl) = 3 mol/L is used as reference electrolyte and storage solution.

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

## Scope of delivery 6.0452.100

| Order Number | Description | Quantity |
|---|---|---|
| 6.1236.050  | **Sleeve with SGJ 14/12 mm**<br>Sleeve with SGJ 14/12 mm, polyethylene. | 1 Stk |
| 6.2008.040  | **Storage vessel**<br>Together with 6.2008.050 storage vessel holder. it provides a support for the electrode on 807 Dosing Units. | 1 Stk |

## Optional accessories

6.2308.020
**Electrolyte 3 mol/L KCl (250 mL)**

Electrolyte solution c(KCl) = 3 mol/L, (for Ag/AgCl reference systems)



---

6.2104.020
**Electrode cable / 1 m / F**

For connecting electrodes with Metrohm plug-in head G to Metrohm instruments (socket F).



---