



KFT equipment for OMNIS Titrator

6.05609.000

Complete accessory set for volumetric Karl Fischer titration.

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 6.05609.000

Order Number	Description	Quantity
6.1903.020	Stirring bar / 16 mm Stirring Bar with magnetic core, PTFE coated, length 16 mm	2 PCS
		
6.1903.030	Stirring bar / 25 mm Stirring bar with magnetic core, PTFE covering, length 25 mm.	2 PCS
		
6.00346.100	Double Pt ring electrode for volumetry Indicator electrode for Karl Fischer titration which, thanks to its two fused platinum rings, is more robust and easier to clean than the conventional platinum pin electrode (6.0338.100).	1 PCS
		

6.01405.010

OMNIS titration vessel lid for volumetric Karl Fischer titration 1 PCS


Titration vessel lid for volumetric Karl Fischer titration with an OMNIS titrator.



6.01406.010

Adsorber tube with lid 1 PCS

Adsorber tube with lid for OMNIS Karl Fischer titrator, volumetric



6.01406.220

20 – 90 mL titration vessel 1 PCS

20 – 90 mL titration vessel with screw top for use with 6014050X0 titration vessel lids.



6.01406.250

50 – 150 mL titration vessel 1 PCS

50 – 150 mL titration vessel with screw top for use with 6014050X0 titration vessel lids.



6.01408.010

Set of seals for Karl Fischer titration cell

1 PCS

Complete set of seals, suitable for 6.05609.000 and 6.05609.010.



6.01543.120

Antidiffusion tip

1 PCS

Titration tip with antidiffusion valve, suitable for OMNIS Karl Fischer titrator.



6.02104.040

Electrode cable plug-in head G (pol.) / plug P, 0.55 m 1 PCS

Cable for connecting an analog polarizable electrode (plug-in head G) to an analog measuring module from OMNIS.



6.02709.010

Stopper

3 PCS

Plastic stoppers, suitable for 2.1001.4220 OMNIS Karl Fischer Titrator.



6.02709.020

Septum stopper

1 PCS

Plastic septum stopper, suitable for 6014050X0 titration vessel lids.



6.02709.030

Screw nipple

2 PCS

Plastic screw nipple, suitable for 6014050X0 titration vessel lids.



6.02801.000

Cotton for adsorber tube

1 PCS

Wad of cotton (4 pieces), suitable for OMNIS 601406010 adsorber tube.



6.1448.010

Septum 12 mm, 5 pieces

2 PCS

Set of 5 items.



6.2412.000


Weighing boats

1 PCS

Weighing boat made of glass for the 6.1414.030 titration vessel lid or the 6.1455.31X, 6.1464.32X, 6.1465.320 titration vessel. For KF water determinations. Including protective tube.

6.2816.020


Syringe 10 mL with Luer connection

2 PCS

For various applications in IC and VA

6.2816.030


Needle with Luer connector

2 PCS

Needle for filling in mercury into the Multi-Mode Electrode and for sample addition in Karl Fischer titration.

Optional accessories

6.02709.040

Stopper



Stopper for titration vessel lids 6014050X0.

6.02711.000

Spoon for paste



Suitable for use in Karl Fischer water content determinations of pasty substances such as honey, cosmetics, greases or various dairy products.

The spoon for paste facilitates the addition of viscous samples. It is handled like any other spatula and enables convenient, rapid, and loss-free addition of pastes into the Karl Fischer titration vessel.

We furthermore recommend the use of a homogenizer for dissolving viscous samples quickly and completely.

Compatible with 6014050X0 titration vessel lids.

6.2811.000

Molecular sieve



Molecular sieve. Bottle containing 250 g. Pore size: 0.3 nm. Without moisture indicator. For Rancimats and Karl Fischer instruments.

6.2811.010

Molecular sieve (US)



Molecular sieve. Bottle containing 250 g. Pore size: 0.3 nm. Without moisture indicator. For Rancimats and Karl Fischer instruments.

This product is intended for all users who are US residents. The material is rated hazardous in the US in accordance with OSHA Hazard Communication Standard (29 CFR 1910.1200).