



863 Compact IC Autosampler

2.863.0010

The 863 Compact IC Autosampler is the ideal help for routine analysis. It enables automation of the full range of Metrohm IC instruments.

Phạm vi giao hàng 2.863.0010

Qt.	Order no.	Description
1 PCS	1.863.0010	863 Compact IC Autosampler

The 863 Compact IC Autosampler is the ideal help for routine analysis. It enables automation of the full range of Metrohm IC instruments.



2 PCS 6.1826.360 Pump tubing LFL (white/white), 3 stoppers
For Sample Processors



1 PCS 6.1831.050 PEEK capillary 0.5 mm i.d., 40 cm
For inline dialysis and for sample changers. For IC Dialysis Unit, IC Sample Processor, Compact IC Autosampler, IC Dialysis Sample Processor, IC Liquid Handling Dialysis Unit, Compact VA Autosampler



1 PCS 6.1831.060 PEEK capillary 0.5 mm i.d., 1 m
For inline dialysis and for Sample Changers. For IC Dialysis Unit, IC Sample Processor, Compact IC Autosampler, IC Filtration Sample Processor, IC Dialysis Sample Processor, IC Liquid Handling Dialysis Unit



1 PCS 6.2151.100 Adapter USB MINI (OTG) - USB A
For connecting USB instruments.



1 PCS

6.2151.110

Metrohm USB Mini B cable (OTG) - USB A, 1.8 m

For connecting USB instruments.



1 PCS

6.2621.090

1/2 in. wrench

For 1/2 in. screws. For IC Sample Processors



1 PCS

6.2621.130

Hexagon key 2 mm

2 mm.



1 PCS

6.2621.140

Hexagon key 2.5 mm



1 PCS

6.2739.000

Wrench

For tightening connectors



1 PCS

6.2743.057

Sample vessel 11 mL

200 pieces. For IC Sample Processors and VA Autosamplers.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



1 PCS

6.2743.077

Stopper with perforation

200 pieces. For IC Sample Processors



1 PCS

6.2744.034

Coupling olive/UNF 10/32, 2x

Connection between pressure screw and pump tubing. 2 pieces. For IC instruments with peristaltic pump.



1 PCS

6.2744.070

Pressure screw short

Short version. With UNF 10/32 connection. 5 pieces. For the connection of PEEK capillaries




1 PCS

6.2744.160

Nozzle/UNF 10/32 coupling with security device

Security device for connecting pump tubing to a nozzle.



1 PCS

6.2833.030

Needle holder 1/8 in. for 863 Compact Autosampler

Needle holder 1/8 in. for 863 Compact Autosampler



1 PCS

6.2846.010

Sample needle made of zirconium oxide with PEEK tip

Sample needle for Sample Processors and Autosampler in ion chromatography. The PEEK tip guarantees secure piercing of the vial covering.



Phụ kiện tùy chọn

Order no.	Description	
6.1826.310	Pump tubing LFL (orange/green), 3-stopper Pump tubing for suppressor solutions and for acceptor solutions for inline dialysis and for inline ultrafiltration.	
6.1826.320	Pump tubing LFL (orange/yellow) , 3-stopper For suppressor solutions and for acceptor solutions for inline dialysis and for inline ultrafiltration.	
6.1826.330	Pump tubing LFL (orange/white), 3 stoppers For all IC instruments with peristaltic pump.	
6.1826.340	Pump tubing LFL (black/black), 3 stoppers For the sample solution in inline dialysis.	

6.1826.380 Pump tubing LFL (gray/gray), 3 stoppers
For Inline Dilution



6.1826.390 Pump tubing LFL (yellow/yellow), 3 stoppers
For the sample solution in Inline Ultrafiltration.



6.2141.230 Remote cable Compact Sample Changer/Compact Autosampler - Titrino plus/Pump Station
Cable for the remote connection between Compact Sample Changer /Compact Autosampler and Titrino plus and/or Pump Station.



6.2141.270 Cable Compact IC Autosampler - Remote Connection on IC Instruments
Enables remote control of the Compact IC Autosampler from IC Instruments



6.2151.020 Cable USB A - USB B 1.8 m

USB connecting cable



6.2629.000 Screw nut to needle holder IC

On Sample Processors



6.2723.310 Protective cover for Titrino plus

Protective cover for display and keypad for all instruments of the Titrino plus family.

Suitable for: 848 Titrino plus, 862 Compact Titrosampler, 863 Compact Autosampler, 865 Dosimat plus, 869 Compact Sample Changer, 870 KF Titrino plus, 876 Dosimat plus, 877 Titrino plus, 899 Coulometer.



6.2743.040 Vial PP, 2.5 mL, 2000 pieces

2.5 mL. 2000 pieces. For IC Sample Processors.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean-room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



6.2743.047

Vial PP, 2.5 mL, 200 pieces

2.5 mL. 200 pieces. For IC Sample Processors.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.



6.2743.050

Sample vessel 11 mL

2000 pieces. For IC Sample Processors and VA Autosamplers.

The IC sample vessels are made of polypropylene (PP) plastic for health-care applications. Production and packaging both take place under sterile and clean room conditions (Clean Room Class 7, in accordance with DIN ISO 14644). This guarantees minimal degrees of contamination. The sample vessels are subjected to random checks for leachable anions and cations.

