



Conductivity measuring cell $c = 0.1 \text{ cm}^{-1}$ with Pt1000 (fixed cable, 0.65 m)

6.00921.040

Conductivity measuring cell made of stainless steel with cell constant $c = 0.1 \text{ cm}^{-1}$ (guide value), with integrated Pt1000 temperature sensor and with fixed cable (0.65 m) for connecting to the OMNIS Measuring Module Conductivity.

This sensor is suitable for measurements of low conductivities, e.g. in deion. water or for measurements in accordance with USP 645>.

Phạm vi giao hàng 6.00921.040

Order Number	Description	Quantity
6.1236.050	Sleeve with SGJ 14/12 mm Sleeve with SGJ 14/12 mm, polyethylene.	1 Stk



Phụ kiện tùy chọn

6.2324.010

**Conductivity standard 100 μ S
/cm 250 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.



6.2324.110

**Conductivity standard 100 μ S
/cm, 5 x 30 mL**

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.



6.02101.020

**Measuring module
conductivity**

Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.

