Conductivity measuring cell c = 0.5 cm-1 with Pt1000 (fixed cable, 2.0 m)

6.00924.080

4-conductor conductivity measuring cell with cell constant c = 0.5 cm^{-1} (guide value), with integrated Pt1000 temperature sensor and with fixed cable (2.0 m) for connecting to the OMNIS Measuring Module Conductivity. Thanks to the robust/break-proof plastic shaft made of PEEK, this sensor is mechanically very resistant and is suitable for automated measurements of medium conductivities, e.g., in:

– Drinking water
– Surface water
– Waste water

Phạm vi giao hàng 6.00924.080

Order Number	Description	Quantity
6.1236.050	**Sleeve with SGJ 14/12 mm**	1 Stk
	Sleeve with SGJ 14/12 mm, polyethylene.	

Phụ kiện tùy chọn

6.2324.010
Conductivity standard 100 µS /cm 250 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.

6.2324.110
Conductivity standard 100 µS /cm, 5 x 30 mL

Conductivity standard for calibration of conductivity measuring cells with cell constant = 0.1/cm.

6.2301.060
KCl ion standard 250 mL

KCl ion standard (c(ion) = 0.1000 +/-0.0005 mol/L), can also be used as standard solution for conductometry 12.87 mS/cm (25 °C).

6.02101.020

Measuring module conductivity

Measurement channel for one OMNIS Titrator or Titration Module for the connection of conductivity measuring cells.