



Application Note AN-S-376

Fluoride in sodium fluoride gel for pharmaceutical use

Method validation according to the U.S. Pharmacopoeia

Fluoride is a mineral that is naturally found in water and some foods. It has been proven to strengthen the enamel of the teeth and protect them from decay [1]. However, exposure to too much fluoride can cause dental fluorosis, a condition that affects the appearance of the teeth. Therefore, it is important to monitor the amount of fluoride in dental care products such as gels and toothpastes.

Sodium fluoride gel is a beneficial product that effectively helps prevent cavities (caries). As specified by the authoritative United States Pharmacopeia – National Formulary (USP-NF) Monograph «Sodium

Fluoride Gel» [2], ion chromatography (IC) with suppressed conductivity detection is a reliable method to measure fluoride and impurities in sodium fluoride gel.

This study validates an IC method using a Metrosep A Supp 16 - 250/4.0 column and a hydroxide eluent, which meets the USP-NF criteria. Fluoride is separated from chloride and other contaminants in gel toothpaste with high accuracy and precision. The IC method has been validated according to USP General Chapters <621> Chromatography [3] and <1225> Validation of Compendial Procedures [4].

Commercial gel toothpaste was diluted to a known concentration of approximately 2 µg/mL sodium fluoride (NaF). Here, 1.585 g of a gel toothpaste containing 331.5 mg NaF/100 g was diluted in 500 mL ultrapure water (UPW). The solution was sonicated for 10 minutes and further diluted 1:8.8 with UPW. Afterwards, the diluted solution was filtered using 0.2 µm pore size filters. The nominal sodium fluoride concentration for these samples was 1.19 µg/mL.

No additional sample preparation is required.

The standard solutions and the system suitability

solutions are prepared from the respective 1000 µg/mL certified standards by dilution with UPW. For the assay, the standard solution is obtained by diluting a sodium fluoride solution to 2 µg/mL. The system suitability solution contains 2 µg/mL sodium fluoride and 1 µg/mL sodium acetate. For the impurity test, the standard solution consists of 0.2 µg/mL sodium chloride in UPW. The system suitability solution for the impurity test contains 1 mg/mL sodium fluoride and 1 µg/mL sodium chloride in UPW.

EXPERIMENTAL

Samples and standard solutions were directly injected

into the IC using a 919 IC Autosampler plus (Figure 1).



Figure 1. Instrumental setup including a 930 Compact IC Flex, 919 IC Autosampler plus, and an 800 Dosino for automatic regeneration of the Metrohm Suppressor Module (MSM).

Fluoride was separated from acetate and chloride by using a potassium hydroxide eluent and the Metrosep A Supp 16 column (column material, Table 1). The

analytes were quantified by evaluating their conductivity signal after chemical suppression. The calibration was performed using a single 2.0

µg/mL sodium fluoride standard injected six times.

The sample was analyzed in duplicate.

Table 1. Requirements for IC method as per USP Monograph «Sodium Fluoride Gel» [2].

Column with L91 packing	Metrosep A Supp 16 - 250/4.0
Eluent	15 mmol/L potassium hydroxide
Flow rate	1.0 mL/min
Temperature	40 °C
Injection volume	20 µL
Detection	Conductivity with suppression

RESULTS

The IC assay for fluoride content was validated according to the USP Monograph «Sodium Fluoride Gel» [2]. Suitability requirements for resolution, tailing

factor, and relative standard deviation were fulfilled (Table 2).

Table 2. Suitability requirements for the fluoride assay.

Parameter (assay)	Actual	USP requirement	Status
Resolution F ⁻ /acetate	5.9	NLT 1.5	Pass
Tailing factor	1.1	NMT 2.0	Pass
RSD fluoride (%, n=5)	0.52	NMT 0.73	Pass

Commercial gel toothpaste samples were analyzed for their sodium fluoride content and the results showed the concentration at 104% of the label claim

(Figure 2). The recovery of fluoride for the sample analysis was within the USP acceptance criteria of 90–110%.

RESULTS




Figure 2. Chromatogram of a commercial toothpaste sample containing 1.24 µg/mL sodium fluoride (104% of the label claim).

When performing the impurity tests for potential contamination with chloride, the IC method showed

excellent compliance with the USP requirements (Table 3).

Table 3. Suitability requirements for the chloride impurity in sodium fluoride gel.

Parameter (impurity)	Actual	USP requirement	Status
Resolution F ⁻ /Cl ⁻	7.7	NLT 4	Pass
RSD fluoride (%), n=5	4.2	NMT 5	Pass
S/N ratio Cl ⁻	>740	NLT 20	Pass

SUMMARY

The presented IC method complies with the USP General Chapters <621> and <1225> [3,4]. It is suitable to determine sodium fluoride in gels

containing sodium fluoride according to the USP Monograph «Sodium Fluoride Gel» [2].

REFERENCES

1. Yeung, C. A. A Systematic Review of the Efficacy and Safety of Fluoridation. *Evid Based Dent* **2008**, *9* (2), 39–43.
<https://doi.org/10.1038/sj.ebd.6400578>.
2. Sodium Fluoride Gel; Monograph; U.S. Pharmacopeia/National Formulary: Rockville, MD.
https://doi.org/10.31003/USPNF_M3947_02_01.
3. 621 Chromatography; General Chapter; U.S. Pharmacopeia/National Formulary: Rockville, MD.
https://doi.org/10.31003/USPNF_M99380_01_01.
4. 1225 Validation of Compendial Procedures; General Chapter; U.S. Pharmacopeia/National Formulary: Rockville, MD.
https://doi.org/10.31003/USPNF_M99945_04_01.

CONTACT

Metrohm Nordic AB
Box 11065
161 11 Bromma

mail@metrohm.se

CONFIGURATION



930 Compact IC Flex Oven/ChS/PP/Deg

The 930 Compact IC Flex Oven/ChS/PP/Deg is the intelligent Compact IC instrument with **column oven**, **chemical suppression** and a **peristaltic pump** for suppressor regeneration and built-in **degasser**. The instrument can be used with any separation and detection methods.

Typical areas of application:

- Anion determinations with chemical suppression and conductivity detection
- Organic acids with ion-exclusion chromatography and inverse suppression

919 IC Autosampler plus

The 919 IC Autosampler plus fulfills the requirements of laboratories with medium sample numbers. It enables automation of the full range of Metrohm IC instruments.



800 Dosino

The 800 Dosino is a drive with write/read hardware for intelligent Dosing Units. With fixed cable (length 150 cm).