

## Application Note AN-S-399

# IC assay for fluoride in sodium fluoride and acidulated phosphate topical solution

# Column equivalency study according to U.S. Pharmacopoeia

Sodium fluoride and acidulated phosphate topical solution products are used for dental hygiene to prevent cavities on teeth. Fluoride is a mineral that helps to strengthen the enamel on teeth, making them more resistant against decay by acids and bacteria.

Pharmaceuticals such as these anticavity products require strict quality control. Therefore, it is necessary to determine any impurities in such products as well as the appropriate contents of the active substances. This Application Note outlines the **fluoride IC assay** as described in the USP Monograph Sodium Fluoride and Acidulated Phosphate Topical Solution [1] using the Metrosep A Supp 1 column and suppressed conductivity detection. The column equivalency study was carried out in cooperation with the U.S. Pharmacopoeia.



#### SAMPLES AND STANDARDS

A dentalcare product (topical solution) with a nominal concentration of 1.23% sodium fluoride was investigated for its fluoride content. The samples were diluted in ultrapure water in two steps to obtain a nominal concentration of  $0.5 \mu g/mL$  fluoride.

The system suitability solution contained 1.1  $\mu$ g/mL sodium fluoride and 0.5  $\mu$ g/mL of sodium chloride in ultrapure water, as described in the respective USP

## EXPERIMENTAL

The samples were injected directly into the ion chromatograph (**Figure 1**) without further sample preparation and analyzed using the method parameters given in the USP monograph (**Table 1**). Anionic components were separated isocratically on a Metrosep A Supp 1 - 250/4.0 column containing the alternative packing material L46. The conductivity signal was detected after sequential suppression.

monograph [1]. The solution was prepared by volumetric dilution of 1000 mg/L standard solutions of fluoride and chloride using micropipettes and 100 mL volumetric flasks.

The quantification of the results was performed based on a single point calibration with a 1.1  $\mu$ g/mL fluoride standard.



**Figure 1.** Instrumental setup including a 930 Compact IC Flex with IC Conductivity Detector and a 919 IC Autosampler plus.

Column with L46 packing	Metrosep A Supp 1 - 250/4.6
Eluent	150 mg/L of anhydrous sodium carbonate and 1.0 mL/L of 1 N sodium hydroxide
Flow rate	1.0 mL/min
Column temp.	30 °C
Injection vol.	20 μL
Detection	Conductivity with sequential suppression

Table 1. IC method parameters as per USP monograph «Sodium Fluoride and Acidulated Phosphate Topical Solution» [1].



#### RESULTS

Fluoride and chloride were analyzed in diluted topical solutions according to the USP General Chapter Chromatography [**2**]. A column equivalency study was performed, and all acceptance criteria (e.g., repeatability, resolution, tailing factor, and accuracy) were fulfilled according to the USP General Chapter Validation of Compendial Procedures [**3**]. Resolution between the fluoride and chloride peaks was exceptional (11.6) using the <u>Metrosep A Supp 1 - 250/4.6</u> column (packing material L46).



Figure 3 shows that the fluoride content was accurately reflected in the declared label for the analyzed samples (99.8% recovery).



Figure 2. Chromatogram of the system suitability solution containing 1.1  $\mu$ g/mL sodium fluoride and 0.5  $\mu$ g/mL sodium chloride.



**Figure 3.** Chromatogram of a topical solution containing 1.065  $\mu$ g/mL fluoride (99.8% recovery) and traces of chloride (not quantified).

Performance characteristics	Acceptance criteria	Resu Its
Resolution	>1.5 between fluoride and chloride peaks	11.6
Tailing factor	Tailing factor (asymmetry) for the fluoride ion peak is <2.0	1.2
Repeatability	Relative standard deviation for the fluoride ion peak area in the standard solution is <2.0% for five replicates	1.2 %

Table 2. Selected performance characteristics from the system suitability study.



#### CONCLUSION

The presented IC method for Fluoride in Sodium Fluoride and Acidulated Phosphate Topical Solution with the Metrosep A Supp 1 column (packing material L46) is officially included into the USP [1]. Robustness and reliability of the method was demonstrated within the column equivalency study following the USP General Chapters for Chromatography and Validation of Compendial Procedures [2–3].

[3]1225 Validation of Compendial Procedures;

General Chapter; U.S. Pharmacopeia/National

Rockville,

ΜD.

DOI:10.31003/USPNF M99380 01 01

DOI:10.31003/USPNF M99945 04 01

Formulary:

AE6-0122-092020

#### REFERENCES

[1] U.S. Pharmacopeia. USP-NF Sodium Fluoride and Acidulated Phosphate Topical Solution. Monograph. <u>https://doi.usp.org/USPNF/USPNF\_M76520\_06\_01.h</u> <u>tml</u>.

[**2**] 621 Chromatography.

Internal references: AW IC AE6-0121-092021; AW IC

#### CONTACT

Metrohm Viet Nam Phan Dinh Giot 70000 Ho Chi Minh

info@metrohm.vn

#### CONFIGURATION



#### 930 Compact IC Flex

The 930 Compact IC Flex is the intelligent Compact IC instrument **without suppression**. The instrument can be used with any separation and detection methods. Typical areas of application:

- Anion or cation determinations without suppression with conductivity detection
- Simple applications with UV/VIS or amperometric detection





#### Metrosep A Supp 1 - 250/4.6

The Metrosep A Supp 1 - 250/4.6 is a universal anion column that is characterized by medium capacity and special selectivity. With this column it is possible to process samples with great differences in concentrations. For example, 4 µg/L of sulfate can be determined in a solution containing 150 g/L sodium chloride. An additional advantage is that bromide does not elute until after the nitrate. Particularly in the area of oxohalide analysis, the A Supp 1 - 250/4.6 impresses with its outstanding separation properties. Pressure fluctuations, constantly changing eluents and large sample throughput do not influence the separating efficiency of this column, even after very long periods. It is the "workhorse" for development and routine laboratories.

