



Application Note AN-T-219

面粉、面和面包中的 pH 和 TTA

Determination of the pH value and the total titratable acidity according to AOAC 943.02, 981.12, and AACC 02-31.01

Bread is one of our most important staple foods. In order to consistently manufacture a high quality product, it is critical to measure certain parameters e.g. pH value or the acidity content in the raw materials and during the production steps. These factors have a major influence on the taste and storage lifetime of the final product.

Many bakers rely on precise weighing of their raw materials, or empirical values. However,

these methods only work to a limited extent. With sourdough for example, the contained lactic acid changes the pH value and the degree of acidity, among other things. These parameters determine factors including taste, aroma, consistency, and shelf life—in short, the quality. Therefore, consistent product quality is only possible with precise measurements during the process.

SAMPLE AND SAMPLE PREPARATION

This application is demonstrated on white flour, bread dough made from white flour, unbaked sourdough made from wheat and rye flour, and

both white and whole wheat bread. No sample preparation is required.

EXPERIMENTAL

The determinations are carried out on an Eco Titrator equipped with an Ecotrode Gel with NTC, a 913 pH Meter, and a Polytron for sample size reduction.

An appropriate amount of sample is weighed into the sample beaker and CO₂-free water is added. If necessary (e.g. for dough or bread), the sample is homogenized with the Polytron and the sample is allowed to stand for 30 minutes.

For determination of the pH value, the supernatant is carefully decanted and the pH is measured immediately with the 913 pH Meter.

For the TTA measurement, the solution is titrated until after the first equivalence point with standardized sodium hydroxide solution is reached.

RESULTS

Well-defined pH values and titration curves are obtained for the tested samples.



Figure 1. Eco Titrator and a 913 pH Meter with a maintenance-free Ecotrode Gel with NTC.

The results are summarized in **Table 1** and **Table 2**. An example titration curve is displayed in **Figure 2**.

Table 1. Results for the pH value according to AOAC 943.02 and AOAC 981.12 with a 913 pH Meter equipped with an Ecotrode Gel with NTC.

Sample (n = 6)	CMean pH value	SD(rel) in %
White flour	6.19	0.0
Bread dough	7.09	0.2
Unbaked sourdough (wheat flour)	5.64	0.3
Unbaked sourdough (rye flour)	4.87	0.1
White bread	5.50	0.2
Whole wheat bread	6.10	0.2

Table 2. Results for the TTA measurement according to AACC 02-31.01 with an Eco Titrator equipped with an Ecotrode Gel with NTC.

Sample (n = 6)	Mean in mL c(NaOH) = 0.1 mol/L per 10 g sample	SD(rel) in %
White flour	2.72	2.2
Bread dough	4.18	0.9
Unbaked sourdough (wheat flour)	5.07	2.2
Unbaked sourdough (rye flour)	6.70	1.5
White bread	4.55	5.5
Whole wheat bread	3.34	4.0




Figure 2. Titration curve of the determination of the TTA of bread dough on an Eco Titrator.

CONCLUSION

Both the Eco Titrator and the 913 pH Meter are unbeatable in combination to determine the reliable key figures pH value and the total titratable acidity. They are low-priced, user-friendly, and take up little

space. Pre-installed methods on the Eco Titrator make it easier for users without laboratory experience to get started with precise and fast titrations, perfect for bakeries.

Internal reference: AW TI CH1-1315-012021

CONTACT

117702
100085

marketing@metrohm.com.cn

CONFIGURATION



Eco Titrator Acid/Base

型 Eco Titrator 具有内置磁力拌器和触摸感式用界面,可以足日常分析需求。其始提供符合 GLP 准的果。

Eco Titrator Acid/Base 提供了用于在水性溶液中酸滴定的完整套件。套件包括了滴定、一个 20 mL 量管元以及一个合式 pH Ecotrode plus。



913 pH Meter

便携式双通道 pH 量,用于量 pH/mV 和温度。通池供且有支架板的量,便可以面向和在室中量的装。

- 内置池和个流隔 pH 量入端的便携式 pH
- 模 pH 量入端,用于万通准 pH
- 数字 pH 量入端,用于万通智能 pH
- 固用的防水防外 (IP67),合外及室使用
- 背光 LCD 彩色示屏,用于方便取果
- USB 接口,用于方便将数据从算机出到打印机上
- 更大的内存(10000 数据)
- 引保型用和家模式,防止修改一些默参数
- GLP 打印和数据出,用 ID 和戳



Pt1000 Ecotrode Gel

免 pH (凝解),有集成温度感器 (Pt1000),用于似品中的常 pH 量。存在 $c(\text{KCl}) = 6.2308.000$,不用于子含量低的溶液。

老化指示器提前示需要更。