



Application Note AN-T-220

# Sodium chloride content in dough and bread

## Determination of the sodium chloride content according to AOAC 971.27

For millenia, bread and salt—the staple food of humankind, has been vital for sustenance. Salt has been known as the «white gold» since ancient times and is what enhances the flavor in many products.

The increase of health consciousness in the past few decades has led to efforts in reducing the sodium chloride content in all food products. If consumed in excess, sodium may damage the cardiovascular system. It is therefore in the interest of food manufacturers to reduce the salt

content and while preserving the flavor of the food.

To ensure consistent quality, it is critical to know the exact salt content in the raw materials and the final products. This is only possible by performing precise measurements during the production process.

It is possible to determine sodium chloride in dough and bread quickly according to AOAC 971.27 with the Eco Titrator equipped with an Ag Titrode.

SAMPLE AND SAMPLE PREPARATION

This application is demonstrated on bread dough made from white flour and whole wheat bread.  
An appropriate amount of sample is weighed

into the sample beaker and CO<sub>2</sub>-free water is added. Then the sample is homogenized with the Polytron.

EXPERIMENTAL

The determinations are carried out on an Eco Titrator equipped with an Ag Titrode and a Polytron for sample preparation.  
An appropriate amount of sample is weighed into the sample beaker and CO<sub>2</sub>-free water as well as nitric acid solution is added.  
While stirring, the solution is titrated until after the first equivalence point with standardized silver nitrate solution.



Figure 1. Eco Titrator equipped with an Ag Titrode.


RESULTS

Well-defined NaCl values and titration curves are obtained for the tested samples.

The results are summarized in Table 1. An exemplary titration curve is displayed in Figure 2.

Table 1. Results for the sodium chloride content according to AOAC 971.27 with an Eco Titrator equipped with an Ag Titrode.

Sample (n = 6)	Mean NaCl in %	SD(rel) in %
Bread dough	1.03	0.1
Whole wheat bread	2.31	0.1



**Figure 2.** Titration curve of the determination of the sodium chloride content of whole wheat bread on an Eco Titrator.

## CONCLUSION

Titration is a precise and reliable method to determine the sodium chloride content in dough and bread.

Using the Eco Titrator equipped with an Ag Titrode allows a fast determination. The system

offers both low-priced and user-friendly handling. Pre-installed methods on the Eco Titrator makes it easy for customers without laboratory experience to get started with precise and fast titrations, perfect for bakeries.

Internal reference: AW TI CH1-1315-012021

## CONTACT

瑞士万通中国  
北京市海淀区上地路1号院  
1号楼7702  
100085 北京

marketing@metrohm.com.cn

## CONFIGURATION



### Eco Titrator Salt

型 Eco Titrator 具有内置磁力拌器和触摸感式用界面,是日常分析的理想。其始提供符合 GLP 准的果,并且具有最小的空需求( DIN A4)。

用于 Eco Titrator 提供用于分析不同品中化物的完整工具包。套件包括了滴定、10 mL 量管元以及一个 Ag-Titrode、一个用于硝酸淀滴定的免。



### Polytron PT 1300 D

Polytron PT 1300 D-Metrohm 版本

均匀,可直接通 OMNIS Software, tiamo™ 或 Touch Control 行控制。

Polytron PT 1300 D 由一个控制装置和一个装置成。通装置上的合系可便快速地更,无需外使用工具。可松粉碎固体品可。也非常用于粘性品的良好混合。