





Accessories for 6.5339.030

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 6.5339.030

Qty.	Order no.	Description
1 pcs	6.0343.100	Separate Pt Rod Electrode Platinum auxiliary electrode for voltammetry and CVS.
		
1 pcs	6.0728.120	Ag/AgCl reference electrode Inner reference system with electrolyte c(KCl) = 3 mol/L. For use with electrolyte vessel made of plastic 6.1245.010.
		
1 pcs	6.1204.500	Stirrer for Professional VA instruments Stirrer for Professional VA instruments, complete, metal-free.
		
1 pcs	6.1226.030	Mercury drop capillary (non-silanized) for MME Non-silanized mercury drop capillary (10 pieces)
		

1 pcs 6.1226.050 Mercury drop capillary (silanized) for MME

Silanized mercury drop capillary (10 pieces)



1 pcs 6.1245.010 Electrolyte vessel made of plastic

Electrolyte vessel made of plastic with ceramic diaphragm for double junction reference electrode.



1 pcs 6.1246.120 Multi-Mode Electrode pro

Mercury electrode for voltammetry. Can be operated as DME, SMDE or HMDE.



1 pcs 6.1247.020 Needles for MME

Needles for Multi Mode Electrode (3 pieces).



1 pcs 6.1415.210 Titration vessel / 10-90 mL



1 pcs 6.2301.100 Pb ion standard 1 g/L 50 mL

Pb ion standard, $c(\text{Pb}^{2+}) = 1.000 \text{ g/L}$



1 pcs **6.2308.020** **Electrolyte 3 mol/L KCl (250 mL)**
Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$, (for Ag/AgCl reference systems)



1 pcs **6.2406.000** **Mercury drop catcher**
Silver wire (to capture mercury drops by amalgamation)



1 pcs **6.2711.030** **Tray for filling the MME**
For 884 Professional VA, 797 VA Computrace, 757 VA Computrace and 647, 663, 694, 747 VA Stands



1 pcs **6.2816.020** **Syringe 10 mL with Luer connection**
For various applications in IC and VA



1 pcs **6.2816.030** **Needle with Luer connector**
Needle for filling in mercury into the Multi-Mode Electrode and for sample addition in Karl Fischer titration.

