






Accessories for 2.757.0110

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 2.757.0110

Qty.	Order no.	Description	
1 pcs	6.0343.000	Separate Pt Rod Electrode Platinum auxiliary electrode for voltammetry and CVS.	
1 pcs	6.0728.020	Ag/AgCl reference electrode Inner reference system with electrolyte $c(\text{KCl}) = 3 \text{ mol/L}$. For use with electrolyte vessel made of plastic 6.1245.010.	
1 pcs	6.1204.090	Stirrer tip Stirrer tip for VA stands	
1 pcs	6.1226.030	Mercury drop capillary (non-silanized) for MME Non-silanized mercury drop capillary (10 pieces)	

1 pcs **6.1244.020** **Driving belt**
For Rotating Disk Electrodes and stirrer in VA Stands (3 pieces).



1 pcs **6.1245.010** **Electrolyte vessel made of plastic**
Electrolyte vessel made of plastic with ceramic diaphragm for double junction reference electrode.



1 pcs **6.1247.020** **Needles for MME**
Needles for Multi Mode Electrode (3 pieces).



1 pcs **6.1415.210** **Titration vessel / 10-90 mL**



7 pcs **6.1446.040** **Threaded stopper / M6**



1 pcs **6.1801.080** **Tubing for nitrogen connection**
For supplying gas to voltammetry instruments



4 pcs **6.1808.000** **Connection bushing M6**



1 pcs **6.1817.000** **Filling tube for MME**
For filling the Multi-Mode-Electrode.



1 pcs **6.1824.000** **Four-way micro dosing tip**

With 4 x M6 thread. For VA measuring stands.



1 pcs **6.2308.020** **Electrolyte 3 mol/L KCl (250 mL)**
Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$, (for Ag/AgCl reference systems)



1 pcs	6.2406.000	Mercury drop catcher Silver wire (to capture mercury drops by amalgamation)	
1 pcs	6.2615.030	Electrode stand for MME Electrode stand for filling and storing the Multi-Mode Electrode.	
1 pcs	6.2703.000	Stand ring for titration vessels	
1 pcs	6.2709.080	Stopper for pipetting opening For 694, 747 VA Stands and VA Computrace	
1 pcs	6.2711.030	Tray for filling the MME For 884 Professional VA, 797 VA Computrace, 757 VA Computrace and 647, 663, 694, 747 VA Stands	
1 pcs	6.2711.040	Drip pan for VA Computrace	

1 pcs 6.2730.030 Stopper

Complete with nipple and O-ring. For 6.1414.030 titration vessel lids. For KF instruments, VA measurement stands, VA Computrace and IC Dilution.



1 pcs 6.2739.000 Wrench

For tightening connectors



1 pcs 6.2816.020 Syringe 10 mL with Luer connection

For various applications in IC and VA



1 pcs 6.2816.030 Needle with Luer connector

Needle for filling in mercury into the Multi-Mode Electrode and for sample addition in Karl Fischer titration.



1 pcs 6.6032.100 757 VA Computrace software CD 2.0

Software for 757 VA Computrace. For the operation systems Windows® 2000 and Windows® XP Professional. Can also be used under Windows® 98.



Optional accessories

Order no.	Description	
2.772.0110	772 Pump Unit Peristaltic pump for direct connection to Sample Processors.	
2.823.0010	823 Membrane Pump Unit Membrane pump for direct connection to Sample Processors.	
6.0728.000	Ag/AgCl reference electrode without electrolyte Inner reference system without electrolyte for use with electrolyte vessel made of glass 6.1245.000.	
6.0728.010	Ag/AgCl reference electrode without electrolyte Inner reference system without electrolyte for use with electrolyte vessel made of plastic 6.1245.010.	
6.1204.130	Silver Electrode Tip Metal electrode tip, silver (Ag). Electrode disk diameter 2 mm. M3 thread. Temperature range: 0 ... 40 °C.	

6.1204.140 Gold Electrode Tip

Metal electrode tip, gold (Au). Electrode disk diameter 2 mm. M3 thread. Temperature range: 0 ... 40 °C.



6.1204.150 Lateral Gold Electrode Tip

Metal electrode tip, gold (Au), with lateral electrode surface. Electrode disk diameter 3 mm. M3 thread. Temperature range: 0 ... 40 °C.



6.1204.170 3 mm platinum electrode tip for CVS

Metal electrode tip made of platinum (Pt), diameter of electrode disk 3 mm, polished surface, M3 thread connection. Temperature range: 0 - 40 °C.



6.1204.180 Ultra Trace electrode tip

Carbon electrode tip from Ultra Trace graphite. Electrode disk diameter 2 mm. M3 thread connection. Temperature range: 0 - 40 °C.



6.1204.210 Driving axle for rotating disk electrode (RDE)

Driving axle for rotating disk electrode (RDE) with titanium axle for VA measuring stands, M3 thread.



6.1204.220 Driving axle for rotating disk electrode (RDE) with mercury contact

Driving axle for rotating disk electrode (RDE) with titanium axle and mercury contact for VA measuring stands, M3 thread.



6.1241.020 Electrode holder

Electrode holder for VA auxiliary electrodes. For use with glassy carbon rod (6.1247.000) or platinum pin (6.1247.010).



6.1245.000 Electrolyte vessel made of glass

Electrolyte vessel made of glass with ground-joint diaphragm (standard ground-joint) for double junction reference electrode.



6.1247.000 Glassy carbon rod (GC)

For electrode holder 6.1241.020 or 6.1241.120 for VA auxiliary electrode.



6.1415.150 Titration vessel / 5-70 mL

6.1415.220 Titration vessel / 20-90 mL

6.1415.250 Titration vessel / 50-150 mL

6.1415.310 **Titration vessel / 70-200 mL**



6.1418.150 **Titration vessel with thermostat jacket / 5-70 mL**



6.1418.220 **Titration vessel with thermostat jacket / 20-90 mL**



6.1418.250 **Titration vessel with thermostat jacket / 50-150 mL**



6.1450.210 **PFA titration vessel / 10 to 90 mL**

For ultratrace analysis and for solutions containing fluoride.



6.1456.210 **Measuring vessel / 10-90 mL**

Measuring vessel for automated VA systems. Dilution vessel for Advanced IC Sample Processor



6.2709.040 Stopper for MME opening on VA Stand

For closing gas tubing of the Multi-Mode Electrode in VA Stands if not in use. E.301.0004 O-rings included.



6.2709.050 Stopper for stirrer opening on VA Stand

E.301.0004 O-rings included. Used for the stirrer opening of VA Stands.



6.5327.000 MVA-Hg, equipment for mercury determination

Complete set of accessories for the determination of mercury as per Application Bulletin 96. Contains rotating gold working electrode, reference electrode, glassy-carbon auxiliary electrode and other accessories.

