





Accessories for 2.898.0110

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Scope of delivery 2.898.0110

Qty.	Order no.	Description	
1 pcs	1.898.0010	XYZ Sample Changer	
2 pcs	6.1236.020	Sleeve with SGJ 14/12 mm Sleeve with SGJ 14/12 mm and O-ring.	
2 pcs	6.1446.000	SGJ stopper / B-14/(15)	
3 pcs	6.1446.010	Stopper Used with Sample Changers	

1 pcs **6.1458.040** **Titration Head, 3 x SGJ 14**
Titration head for 120 mL beakers. With 3 openings SGJ 14 and 4 openings 6.4 mm, 2 dosing tips included. Suitable for rinsing and aspiration equipment.



1 pcs **6.1459.300** **Sample beaker, plastic (PP), 120 mL, 250 pieces**
Sample beaker, plastic (polypropylene, PP), 120 mL, suitable for sample racks: 6.2041.470, 6.2041.840, 6.2041.850, 6.2068.010, 6.2068.020, 6.2068.030, and 6.9920.141, 250 pieces



2 pcs **6.1621.010** **PE canister / 30 L**
Eluent canister for Online-IC instruments.



2 pcs **6.1621.110** **Cap for 30-L canister**
Cap for 30-L canister. For easy and secure transportation of filled canisters.



2 pcs **6.1805.530** **FEP tubing / M6 / 2 m**
With light and kink protection



1 pcs 6.2071.020 Sample rack for 898 XYZ SC 41 x 120 mL left

Sample rack for 41 x 120 mL beakers (6.1459.300) for the 898 XYZ Sample Changer. The sample rack has a recess for the external rinsing station and is numbered from 1 - 41. It is positioned on the left half of the sample changer. There is sufficient space for a maximum of two sample racks per 898 XYZ SC.



1 pcs 6.2071.030 Sample rack for 898 XYZ SC 41 x 120 mL right







Sample rack for 41 x 120 mL beakers (6.1459.300) for the 898 XYZ Sample Changer. The sample rack has a recess for the external rinsing station and is numbered from 42 - 82. It is positioned on the right half of the sample changer. There is sufficient space for a maximum of two sample racks per 898 XYZ SC.



1 pcs 6.6067.000 USB Stick for 898 XYZ Sample Changer



Optional accessories

Order no.	Description	
6.2071.040	Sample rack for 898 XYZ SC 24 x 250 mL left	
	Sample rack for 24 x 250 mL beakers (6.1432.320) for the 898 XYZ Sample Changer. The sample rack has a recess for the external rinsing station and is numbered from 1 - 24. It is positioned on the left half of the sample changer. There is sufficient space for a maximum of two sample racks per 898 XYZ SC.	
6.2071.050	Sample rack for 898 XYZ SC 24 x 250 mL right	
	Sample rack for 24 x 250 mL beakers (6.1432.320) for the 898 XYZ Sample Changer. The sample rack has a recess for the external rinsing station and is numbered from 25 - 48. It is positioned on the right half of the sample changer. There is sufficient space for a maximum of two sample racks per 898 XYZ SC.	
6.7600.010	898 XYZ Sample Changer drip pan	
	Drip pan for 898 XYZ Sample Changer.	
6.7600.020	898 XYZ Sample Changer rinsing station holder	
	Rinsing station holder for 898 XYZ Sample Changer.	
6.7600.030	898 XYZ Sample Changer rinsing station	
	Rinsing station for 898 XYZ Sample Changer.	
6.7600.040	898 XYZ Sample Changer tubing guide	
	898 XYZ Sample Changer tubing guide	

6.7600.050 898 XYZ Sampler Changer titration head holder

898 XYZ Sampler Changer titration head holder



6.7600.060 898 XYZ Sample Changer safety shield

898 XYZ Sample Changer safety shield



6.7600.070 898 XYZ Sample Changer rinse and extraction kit

898 XYZ Sample Changer rinse and extraction kit (FEP/TYGON).



6.7600.080 898 XYZ Sample Changer rinse reduction kit

898 XYZ Sample Changer rinse reduction kit.



6.7600.090 898 XYZ Sample Changer 24 volt power supply unit

898 XYZ Sample Changer 24 volt power supply unit.

